

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP081425

Review By	Rahul	Review On	8/15/2025 9:35:16 AM
Supervise By	Jagrut	Supervise On	8/15/2025 3:48:12 PM
SubDirectory	BP081425	HP Acquire Method	BNA_P
		HP Processing Method	bp081425

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6757 SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6836 S13170,10ul/1000ul sample SP6770

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025390.D	14 Aug 2025 10:30		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025391.D	14 Aug 2025 11:52	A Fresh Calibration is required.	CG/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BP025392.D	14 Aug 2025 12:33		CG/JU	Ok
4	SSTDICC005	SSTDICC005	BP025393.D	14 Aug 2025 13:15		CG/JU	Ok,M
5	SSTDICC010	SSTDICC010	BP025394.D	14 Aug 2025 13:56		CG/JU	Ok,M
6	SSTDICC020	SSTDICC020	BP025395.D	14 Aug 2025 14:38		CG/JU	Ok,M
7	SSTDICCC040	SSTDICCC040	BP025396.D	14 Aug 2025 15:19	Benzidine failed in the Calibration Method.	CG/JU	Ok,M
8	SSTDICC050	SSTDICC050	BP025397.D	14 Aug 2025 16:01		CG/JU	Ok,M
9	SSTDICC060	SSTDICC060	BP025398.D	14 Aug 2025 16:42		CG/JU	Ok
10	SSTDICC080	SSTDICC080	BP025399.D	14 Aug 2025 17:24		CG/JU	Ok,M
11	SSTDICV040	ICVBP081425	BP025400.D	14 Aug 2025 18:05		CG/JU	Ok,M
12	PB169200TB	PB169200TB	BP025401.D	14 Aug 2025 19:28		CG/JU	Ok
13	PB169210BS	PB169210BS	BP025402.D	14 Aug 2025 20:10		CG/JU	Ok,M
14	SSTDCCC040	SSTDCCC040EC	BP025403.D	14 Aug 2025 20:51		CG/JU	Ok,M
15	DFTPP	DFTPP	BP025404.D	14 Aug 2025 22:14		CG/JU	Ok
16	SSTDCCC040	SSTDCCC040	BP025405.D	14 Aug 2025 22:56		CG/JU	Ok,M
17	PB169228BL	PB169228BL	BP025406.D	14 Aug 2025 23:37		CG/JU	Ok

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP081425

Review By	Rahul	Review On	8/15/2025 9:35:16 AM		
Supervise By	Jagrut	Supervise On	8/15/2025 3:48:12 PM		
SubDirectory	BP081425	HP Acquire Method	BNA_P	HP Processing Method	bp081425
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S13170,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	PB169228BS	PB169228BS	BP025407.D	15 Aug 2025 00:18		CG/JU	Ok,M
19	Q2820-10	SOIL-DUP-2	BP025408.D	15 Aug 2025 01:00		CG/JU	Ok
20	Q2820-11	10PC-W	BP025409.D	15 Aug 2025 01:41		CG/JU	Ok
21	Q2820-21	22M-N	BP025410.D	15 Aug 2025 02:22	Surrogate Fail	CG/JU	ReRun
22	Q2820-22	22M-W	BP025411.D	15 Aug 2025 03:03	Surrogate Fail	CG/JU	ReRun
23	Q2820-14	10P-E	BP025412.D	15 Aug 2025 03:45		CG/JU	Ok
24	Q2820-16	10P-N	BP025413.D	15 Aug 2025 04:26		CG/JU	Ok
25	Q2820-17	88H-E	BP025414.D	15 Aug 2025 05:07		CG/JU	Ok
26	Q2820-17MS	88H-EMS	BP025415.D	15 Aug 2025 05:48		CG/JU	Ok,M
27	Q2820-17MSD	88H-EMSD	BP025416.D	15 Aug 2025 06:29		CG/JU	Ok,M
28	Q2820-18	88H-N	BP025417.D	15 Aug 2025 07:10		CG/JU	Ok
29	Q2820-19	88H-W	BP025418.D	15 Aug 2025 07:51	Surrogate fail	CG/JU	ReRun
30	Q2820-20	88H-S	BP025419.D	15 Aug 2025 08:32	Surrogate Fail	CG/JU	ReRun
31	SSTDCCC040	SSTDCCC040EC	BP025420.D	15 Aug 2025 09:13		CG/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP082525

Review By	Rahul	Review On	8/26/2025 10:26:59 AM
Supervise By	Jagrut	Supervise On	8/26/2025 12:54:30 PM
SubDirectory	BP082525	HP Acquire Method	BNA_P
		HP Processing Method	bp081425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13173,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025560.D	25 Aug 2025 10:24		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025561.D	25 Aug 2025 12:13		CG/JU	Ok
3	PB169362BL	PB169362BL	BP025562.D	25 Aug 2025 12:54		CG/JU	Ok
4	SP6865	SP6865	BP025563.D	25 Aug 2025 13:36		CG/JU	Ok
5	Q2933-04	29-SEPTIC	BP025564.D	25 Aug 2025 14:21		CG/JU	Ok
6	Q2933-04MS	29-SEPTICMS	BP025565.D	25 Aug 2025 15:02		CG/JU	Ok
7	Q2933-04MSD	29-SEPTICMSD	BP025566.D	25 Aug 2025 15:43		CG/JU	Ok
8	Q2932-05	ARS20-0002-CONCRE	BP025567.D	25 Aug 2025 16:25		CG/JU	Ok
9	Q2933-02	25-SEPTIC	BP025568.D	25 Aug 2025 17:06		CG/JU	Ok
10	Q2909-02	4065	BP025569.D	25 Aug 2025 17:48	Internal Standard Fail	CG/JU	Ok,M
11	Q2909-02MS	4065MS	BP025570.D	25 Aug 2025 18:29	Internal Standard Fail	CG/JU	Ok,M
12	Q2909-02MSD	4065MSD	BP025571.D	25 Aug 2025 19:11	Internal standard fail	CG/JU	Ok,M
13	Q2932-01	ARS20-0009	BP025572.D	25 Aug 2025 19:52		CG/JU	Ok
14	Q2949-01	367	BP025573.D	25 Aug 2025 20:33		CG/JU	Ok
15	Q2941-01	SOIL-COMP(1-2)	BP025574.D	25 Aug 2025 21:14		CG/JU	Ok,M
16	Q2947-01	291	BP025575.D	25 Aug 2025 21:55		CG/JU	Ok,M
17	Q2932-03	ARS20-0002-SOIL	BP025576.D	25 Aug 2025 22:37	Out of tune time	CG/JU	Not Ok
18	Q2947-03	235	BP025577.D	25 Aug 2025 23:18	Out of tune time	CG/JU	Not Ok

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP082525

Review By	Rahul	Review On	8/26/2025 10:26:59 AM		
Supervise By	Jagrut	Supervise On	8/26/2025 12:54:30 PM		
SubDirectory	BP082525	HP Acquire Method	BNA_P	HP Processing Method	bp081425
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S13173,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP082625

Review By	Rahul	Review On	8/27/2025 9:27:24 AM
Supervise By	Jagrut	Supervise On	8/27/2025 11:38:38 AM
SubDirectory	BP082625	HP Acquire Method	BNA_P
		HP Processing Method	bp081425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13173,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025578.D	26 Aug 2025 08:59		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025579.D	26 Aug 2025 10:21		CG/JU	Ok
3	PB169325TB	PB169325TB	BP025580.D	26 Aug 2025 11:02		CG/JU	Ok
4	PB169362BS	PB169362BS	BP025581.D	26 Aug 2025 11:44		CG/JU	Ok
5	Q2941-02	SOIL-COMP(1-2)	BP025582.D	26 Aug 2025 12:29		CG/JU	Ok
6	Q2924-05DL	MW-2B-082025DL	BP025583.D	26 Aug 2025 13:10	Need 2X further Dilution	CG/JU	Dilution
7	Q2932-03	ARS20-0002-SOIL	BP025584.D	26 Aug 2025 13:51	Internal Standard Fail	CG/JU	Ok,M
8	Q2926-01	TR-04-08212025	BP025585.D	26 Aug 2025 14:33		CG/JU	Ok
9	Q2924-05DL2	MW-2B-082025DL2	BP025586.D	26 Aug 2025 15:14		CG/JU	Ok
10	Q2940-01	COMP(1-6)	BP025587.D	26 Aug 2025 15:55		CG/JU	Ok
11	Q2924-08	MW-16A-082025	BP025588.D	26 Aug 2025 16:37		CG/JU	Ok
12	Q2924-09	MW-16C-082025	BP025589.D	26 Aug 2025 17:18		CG/JU	Ok
13	Q2924-06	MW-2A-082025	BP025590.D	26 Aug 2025 17:59		CG/JU	Ok
14	Q2936-02	MW-19B-082125	BP025591.D	26 Aug 2025 18:41	Need 20X Dilution, Then decide further dilution.	CG/JU	Dilution
15	Q2947-03	235	BP025592.D	26 Aug 2025 19:22		CG/JU	Ok,M
16	Q2892-02	1-FLOOR-NORTH-EAS	BP025593.D	26 Aug 2025 20:03		CG/JU	Ok
17	Q2928-01	27-BELL	BP025594.D	26 Aug 2025 20:44		CG/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP082725

Review By	Rahul	Review On	8/28/2025 8:32:26 AM
Supervise By	Jagrut	Supervise On	8/28/2025 1:40:30 PM
SubDirectory	BP082725	HP Acquire Method	BNA_P
		HP Processing Method	bp081425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13173,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025595.D	27 Aug 2025 08:58		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025596.D	27 Aug 2025 09:39		CG/JU	Ok
3	PB169370BL	PB169370BL	BP025597.D	27 Aug 2025 10:20		CG/JU	Ok
4	PB169370BS	PB169370BS	BP025598.D	27 Aug 2025 11:01		CG/JU	Ok
5	PB169382BL	PB169382BL	BP025599.D	27 Aug 2025 11:42		CG/JU	Ok
6	PB169382BS	PB169382BS	BP025600.D	27 Aug 2025 12:23		CG/JU	Ok
7	Q2924-02	MW-201-082025	BP025601.D	27 Aug 2025 13:37	Need 20X dilution then decide further dilution	CG/JU	Dilution
8	Q2924-03MS	MW-201-082025MS	BP025602.D	27 Aug 2025 14:18		CG/JU	Ok,M
9	Q2924-04MSD	MW-201-082025MSD	BP025603.D	27 Aug 2025 15:00		CG/JU	Ok,M
10	Q2936-02DL	MW-19B-082125DL	BP025604.D	27 Aug 2025 15:41	Need further 5X dilution	CG/JU	Dilution
11	Q2936-02DL2	MW-19B-082125DL2	BP025605.D	27 Aug 2025 16:22		CG/JU	Ok
12	Q2921-01	MW1	BP025606.D	27 Aug 2025 17:03		CG/JU	Ok
13	Q2924-02DL	MW-201-082025DL	BP025607.D	27 Aug 2025 17:44	Need further 10X dilution	CG/JU	Dilution
14	Q2924-02DL2	MW-201-082025DL2	BP025608.D	27 Aug 2025 18:26		CG/JU	Ok

M : Manual Integration