

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2921

Client ID: SSTDCCC040

Date Analyzed: 08/25/2025

Lab File ID: BP025561.D

Time Analyzed: 12:13

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	900532	17.189	945335	21.63	1072630	25.001
UPPER LIMIT	1801060	17.689	1890670	22.13	2145260	25.501
LOWER LIMIT	450266	16.689	472668	21.13	536315	24.501
EPA SAMPLE NO.						
PB169362BL	890876	17.19	1139400	21.63	1287990	25.00

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.