

Report of Analysis

Client:	G Environmental		Date Collected:	08/20/25	
Project:	120 Petro NJ		Date Received:	08/20/25	
Client Sample ID:	MW1		SDG No.:	Q2921	
Lab Sample ID:	Q2921-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087659.D	1	08/25/25 15:43	VN082525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	4.50	J	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	1.80		0.14	1.00	ug/L

Report of Analysis

Client:	G Environmental	Date Collected:	08/20/25
Project:	120 Petro NJ	Date Received:	08/20/25
Client Sample ID:	MW1	SDG No.:	Q2921
Lab Sample ID:	Q2921-01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087659.D	1	08/25/25 15:43	VN082525

[illegible]

Report of Analysis

Client:	G Environmental	Date Collected:	08/20/25
Project:	120 Petro NJ	Date Received:	08/20/25
Client Sample ID:	MW1	SDG No.:	Q2921
Lab Sample ID:	Q2921-01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087659.D	1	08/25/25 15:43	VN082525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
103-65-1	n-propylbenzene	0.36	J		13.0	ug/L
108-67-8	1,3,5-Trimethylbenzene	2.20	J		13.2	ug/L
95-63-6	1,2,4-Trimethylbenzene	2.90	J		13.5	ug/L
002039-89-6	Benzene, 2-ethenyl-1,4-dimethyl-	20.7	J		15.0	ug/L
000119-64-2	Naphthalene, 1,2,3,4-tetrahydro-	7.60	J		15.2	ug/L
017057-82-8	1H-Indene, 2,3-dihydro-1,2-dimethy	5.90	J		15.3	ug/L
020836-11-7	2,2-Dimethylindene, 2,3-dihydro-	8.50	J		15.4	ug/L
017301-32-5	Undecane, 4,7-dimethyl-	5.20	J		15.8	ug/L
006682-71-9	1H-Indene, 2,3-dihydro-4,7-dimethy	5.60	J		15.9	ug/L
1010189-31-0	2,3,4,5,6,7-Hexahydro-1H-cyclopent	8.70	J		16.1	ug/L
002809-64-5	Naphthalene, 1,2,3,4-tetrahydro-5-	11.9	J		16.5	ug/L
004175-54-6	Naphthalene, 1,2,3,4-tetrahydro-1,	6.30	J		16.7	ug/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products