

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
Lab Code: ACE SDG NO.: Q2921
Lab File ID: VN087647.D Date Analyzed: 08/25/2025
Instrument ID: MSVOA_N Time Analyzed: 09:37
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	257927	8.21	497405	9.08	459274	11.85
UPPER LIMIT	515854	8.706	994810	9.583	918548	12.347
LOWER LIMIT	128964	7.706	248703	8.583	229637	11.347
EPA SAMPLE NO.						
MW1	169994	8.21	395366	9.08	375883	11.85
VN0825WBL01	190267	8.21	429125	9.08	403882	11.84
VN0825WBS01	223180	8.21	452330	9.08	403500	11.84
VN0825WBSD01	230804	8.21	442728	9.08	414378	11.85

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.