

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2921
 Lab File ID: VN087647.D Date Analyzed: 08/25/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:37
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	238788	13.771				
UPPER LIMIT	477576	14.271				
LOWER LIMIT	119394	13.271				
EPA SAMPLE NO.						
MW1	178142	13.77				
VN0825WBL01	182604	13.77				
VN0825WBS01	202288	13.77				
VN0825WBSD01	204092	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.