

## **DATA PACKAGE**

GC SEMI-VOLATILES

**PROJECT NAME : SCA-K262**

**ATC GROUP SERVICES LLC**

**104 East 25th Street**

**New York, NY - 10010**

**Phone No: 212-353-8280**

**ORDER ID : Q2922**

**ATTENTION : Denise Cosenza**



**Laboratory Certification ID # 20012**



|                                 |    |
|---------------------------------|----|
| 1) Signature Page               | 3  |
| 2) Case Narrative               | 4  |
| 2.1) PCB Group1- Case Narrative | 4  |
| 3) Qualifier Page               | 6  |
| 4) QA Checklist                 | 7  |
| 5) PCB Group1 Data              | 8  |
| 6) Shipping Document            | 30 |
| 6.1) CHAIN OF CUSTODY           | 31 |
| 6.2) ROC                        | 33 |
| 6.3) Lab Certificate            | 39 |

|   |
|---|
| 1 |
| 2 |
| 3 |
| 4 |
| 5 |
| 6 |

## Cover Page

**Order ID :** Q2922

**Project ID :** SCA-K262

**Client :** ATC Group Services LLC

### Lab Sample Number

Q2922-02  
Q2922-03  
Q2922-05  
Q2922-06  
Q2922-08  
Q2922-09  
Q2922-11  
Q2922-12  
Q2922-14  
Q2922-15  
Q2922-17  
Q2922-18  
Q2922-20

### Client Sample Number

K262-16A  
K262-16B  
K262-17A  
K262-17B  
K262-18A  
K262-18B  
K262-19A  
K262-19B  
K262-20A  
K262-20B  
K262-21A  
K262-21B  
K262-22A

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : \_\_\_\_\_

Date: 9/10/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

## **CASE NARRATIVE**

**ATC Group Services LLC**

**Project Name: SCA-K262**

**Project # N/A**

**Order ID # Q2922**

**Test Name: PCB Group1**

### **A. Number of Samples and Date of Receipt:**

13 Solid samples were received on 08/20/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: PCB Group1. This data package contains results for PCB Group1.

### **C. Analytical Techniques:**

The analyses were performed on instrument GCECD\_P. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11. The analyses were performed on instrument GCECD\_O. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11. The analysis of PCB Group1s was based on method 8082A and extraction was done based on method 3541.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Retention Times were met for all analysis.

The MS {Q2914-13MS} with File ID: PO113170.D recoveries met the requirements for all compounds except for [AR1016(1)1010% - AR1016(2)1063%], [AR1260(1)638% - AR1260(2)532%] due to matrix interference.

The MS {Q2923-01MS} with File ID: PP074825.D recoveries met the requirements for all compounds except for [AR1016(1)213% - AR1016(2)88%], [AR1260(1)127% - AR1260(2)92%] due to matrix interference.

The MSD {Q2914-13MSD} with File ID: PO113171.D recoveries met the requirements for all compounds except for [AR1016(1)1169% - AR1016(2)1169%], [AR1260(1)691% - AR1260(2)531%] due to matrix interference.

The MSD {Q2923-01MSD} with File ID: PP074826.D recoveries met the requirements for all compounds except for [AR1016(1)206% - AR1016(2)93%], [AR1260(1)126% - AR1260(2)86%] due to matrix interference.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID PO113428.D met the requirements except for Decachlorobiphenyl is failing in 1st column, however it is passed in 2nd column therefore no corrective action was taken AND Aroclor-1260(Peak-03) is failing in 2nd column however it is passed in 1st column therefore no corrective action was taken.

The Continuous Calibration File ID PP074804.D met the requirements except for Aroclor-1260(Peak-03),Aroclor-1260(Peak-04),Aroclor-1260(Peak-05),Decachlorobiphenyl is failing in 2nd column however it is passed in 1st column therefore no corrective action was taken.

Samples K262-19A, K262-20A, K262-21A and K262-22A were diluted due to high concentrations.

#### **E. Additional Comments:**

The soil samples results are based on a dry weight basis.

#### **F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_

## DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

|           |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |
|-----------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Value     | If the result is a value greater than or equal to the detection limit, report the value                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                        |
| <b>U</b>  | Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.                                                                                                                                                                                                                                                                                      |
| <b>ND</b> | Indicates the analyte was analyzed for, but not detected                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                       |
| <b>J</b>  | Indicates an estimated value. This flag is used: <ul style="list-style-type: none"> <li>(1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.)</li> <li>(2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.</li> </ul> |
| <b>B</b>  | Indicates the analyte was found in the blank as well as the sample report as “12 B”.                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |
| <b>E</b>  | Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.                                                                                                                                                                                                                                                                                                                                                                                                                                                                              |
| <b>D</b>  | This flag identifies all compounds identified in an analysis at a secondary dilution factor.                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                   |
| <b>P</b>  | This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.                                                                                                                                                                                                                                                                                                                                                                             |
| <b>N</b>  | This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.                                                                                                                                                                                                                                                                       |
| <b>A</b>  | This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.                                                                                                                                                                                                                                                                                                                                                                                                                                                                                          |
| <b>Q</b>  | Indicates the LCS did not meet the control limits requirements                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                 |

## APPENDIX A

### QA REVIEW GENERAL DOCUMENTATION

Project #: Q2922

Completed

For thorough review, the report must have the following:

#### GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

#### COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

#### CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

#### ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 09/10/2025

### Hit Summary Sheet SW-846

SDG No.: Q2922

Order ID: Q2922

Client: ATC Group Services LLC

Project ID: SCA-K262

| Sample ID              | Client ID  | Matrix | Parameter    | Concentration | C    | MDL | RDL  | Units |
|------------------------|------------|--------|--------------|---------------|------|-----|------|-------|
| Client ID : K262-16A   |            |        |              |               |      |     |      |       |
| Q2922-02               | K262-16A   | SOIL   | Aroclor-1254 | 133           | 3.60 |     | 19.3 | ug/kg |
| Total Concentration:   |            |        |              | 133.000       |      |     |      |       |
| Client ID : K262-16B   |            |        |              |               |      |     |      |       |
| Q2922-03               | K262-16B   | SOIL   | Aroclor-1254 | 48.9          | 3.80 |     | 19.9 | ug/kg |
| Total Concentration:   |            |        |              | 48.900        |      |     |      |       |
| Client ID : K262-17A   |            |        |              |               |      |     |      |       |
| Q2922-05               | K262-17A   | SOIL   | Aroclor-1254 | 155           | 3.60 |     | 19.1 | ug/kg |
| Total Concentration:   |            |        |              | 155.000       |      |     |      |       |
| Client ID : K262-17B   |            |        |              |               |      |     |      |       |
| Q2922-06               | K262-17B   | SOIL   | Aroclor-1254 | 87.8          | 3.60 |     | 19.1 | ug/kg |
| Total Concentration:   |            |        |              | 87.800        |      |     |      |       |
| Client ID : K262-18A   |            |        |              |               |      |     |      |       |
| Q2922-08               | K262-18A   | SOIL   | Aroclor-1254 | 150           | 3.60 |     | 19.2 | ug/kg |
| Total Concentration:   |            |        |              | 150.000       |      |     |      |       |
| Client ID : K262-18B   |            |        |              |               |      |     |      |       |
| Q2922-09               | K262-18B   | SOIL   | Aroclor-1254 | 93.8          | 3.70 |     | 19.5 | ug/kg |
| Total Concentration:   |            |        |              | 93.800        |      |     |      |       |
| Client ID : K262-19A   |            |        |              |               |      |     |      |       |
| Q2922-11               | K262-19A   | SOIL   | Aroclor-1254 | 2000 E        | 3.70 |     | 19.6 | ug/kg |
| Total Concentration:   |            |        |              | 2,000.000     |      |     |      |       |
| Client ID : K262-19ADL |            |        |              |               |      |     |      |       |
| Q2922-11DL             | K262-19ADL | SOIL   | Aroclor-1254 | 2000 D        | 37.0 |     | 196  | ug/kg |
| Total Concentration:   |            |        |              | 2,000.000     |      |     |      |       |



### Hit Summary Sheet SW-846

SDG No.: Q2922

Order ID: Q2922

Client: ATC Group Services LLC

Project ID: SCA-K262

| Sample ID              | Client ID  | Matrix | Parameter    | Concentration | C | MDL  | RDL  | Units |
|------------------------|------------|--------|--------------|---------------|---|------|------|-------|
| Client ID : K262-19B   |            |        |              |               |   |      |      |       |
| Q2922-12               | K262-19B   | SOIL   | Aroclor-1254 | 88.0          |   | 3.50 | 18.7 | ug/kg |
| Total Concentration:   |            |        |              | 88.000        |   |      |      |       |
| Client ID : K262-20A   |            |        |              |               |   |      |      |       |
| Q2922-14               | K262-20A   | SOIL   | Aroclor-1254 | 374 E         |   | 3.50 | 18.3 | ug/kg |
| Total Concentration:   |            |        |              | 374.000       |   |      |      |       |
| Client ID : K262-20ADL |            |        |              |               |   |      |      |       |
| Q2922-14DL             | K262-20ADL | SOIL   | Aroclor-1254 | 372 D         |   | 6.90 | 36.7 | ug/kg |
| Total Concentration:   |            |        |              | 372.000       |   |      |      |       |
| Client ID : K262-21A   |            |        |              |               |   |      |      |       |
| Q2922-17               | K262-21A   | SOIL   | Aroclor-1254 | 936 E         |   | 3.40 | 18.2 | ug/kg |
| Total Concentration:   |            |        |              | 936.000       |   |      |      |       |
| Client ID : K262-21ADL |            |        |              |               |   |      |      |       |
| Q2922-17DL             | K262-21ADL | SOIL   | Aroclor-1254 | 821 D         |   | 17.1 | 90.8 | ug/kg |
| Total Concentration:   |            |        |              | 821.000       |   |      |      |       |
| Client ID : K262-22A   |            |        |              |               |   |      |      |       |
| Q2922-20               | K262-22A   | SOIL   | Aroclor-1254 | 1500 E        |   | 3.70 | 19.5 | ug/kg |
| Total Concentration:   |            |        |              | 1,500.000     |   |      |      |       |
| Client ID : K262-22ADL |            |        |              |               |   |      |      |       |
| Q2922-20DL             | K262-22ADL | SOIL   | Aroclor-1254 | 1300 D        |   | 36.8 | 195  | ug/kg |
| Total Concentration:   |            |        |              | 1,300.000     |   |      |      |       |



# SAMPLE DATA

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-16A               |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-02               |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 88.1       | Decanted: |
| Sample Wt/Vol:     | 30.04                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PO113145.D        | 1         | 08/21/25 09:35 | 08/21/25 20:13 | PB169337      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 4.50  | U         | 4.50     | 19.3       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 4.60  | U         | 4.60     | 19.3       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 4.20  | U         | 4.20     | 19.3       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 4.50  | U         | 4.50     | 19.3       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 6.70  | U         | 6.70     | 19.3       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 133   |           | 3.60     | 19.3       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 5.70  | U         | 5.70     | 19.3       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 4.10  | U         | 4.10     | 19.3       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 3.70  | U         | 3.70     | 19.3       | ug/kg             |
| Total PCBs        | Total PCBs           | 133   |           | 3.60     | 19.3       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 21.4  |           | 32 - 144 | 107%       | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 28.5  |           | 32 - 175 | 142%       | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-16B               |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-03               |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 85.2       | Decanted: |
| Sample Wt/Vol:     | 30.06                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PO113438.D        | 1         | 09/03/25 09:10 | 09/03/25 17:27 | PB169519      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 4.60  | U         | 4.60     | 19.9       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 4.70  | U         | 4.70     | 19.9       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 4.40  | U         | 4.40     | 19.9       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 4.70  | U         | 4.70     | 19.9       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 6.90  | U         | 6.90     | 19.9       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 48.9  |           | 3.80     | 19.9       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 5.90  | U         | 5.90     | 19.9       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 4.20  | U         | 4.20     | 19.9       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 3.80  | U         | 3.80     | 19.9       | ug/kg             |
| Total PCBs        | Total PCBs           | 48.9  |           | 3.80     | 19.9       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 15.2  |           | 32 - 144 | 76%        | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 13.5  |           | 32 - 175 | 68%        | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-17A               |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-05               |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 88.9       | Decanted: |
| Sample Wt/Vol:     | 30.01                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PO113146.D        | 1         | 08/21/25 09:35 | 08/21/25 20:31 | PB169337      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 4.40  | U         | 4.40     | 19.1       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 4.50  | U         | 4.50     | 19.1       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 4.20  | U         | 4.20     | 19.1       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 4.50  | U         | 4.50     | 19.1       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 6.70  | U         | 6.70     | 19.1       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 155   |           | 3.60     | 19.1       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 5.60  | U         | 5.60     | 19.1       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 4.00  | U         | 4.00     | 19.1       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 3.60  | U         | 3.60     | 19.1       | ug/kg             |
| Total PCBs        | Total PCBs           | 155   |           | 3.60     | 19.1       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 22.2  |           | 32 - 144 | 111%       | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 26.9  |           | 32 - 175 | 134%       | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-17B               |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-06               |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 88.9       | Decanted: |
| Sample Wt/Vol:     | 30.05                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PO113439.D        | 1         | 09/03/25 09:10 | 09/03/25 17:44 | PB169519      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 4.40  | U         | 4.40     | 19.1       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 4.50  | U         | 4.50     | 19.1       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 4.20  | U         | 4.20     | 19.1       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 4.50  | U         | 4.50     | 19.1       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 6.60  | U         | 6.60     | 19.1       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 87.8  |           | 3.60     | 19.1       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 5.60  | U         | 5.60     | 19.1       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 4.00  | U         | 4.00     | 19.1       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 3.60  | U         | 3.60     | 19.1       | ug/kg             |
| Total PCBs        | Total PCBs           | 87.8  |           | 3.60     | 19.1       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 19.3  |           | 32 - 144 | 97%        | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 20.7  |           | 32 - 175 | 104%       | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-18A               |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-08               |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 88.2       | Decanted: |
| Sample Wt/Vol:     | 30.05                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PO113147.D        | 1         | 08/21/25 09:35 | 08/21/25 20:49 | PB169337      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 4.50  | U         | 4.50     | 19.2       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 4.60  | U         | 4.60     | 19.2       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 4.20  | U         | 4.20     | 19.2       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 4.50  | U         | 4.50     | 19.2       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 6.70  | U         | 6.70     | 19.2       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 150   |           | 3.60     | 19.2       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 5.70  | U         | 5.70     | 19.2       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 4.10  | U         | 4.10     | 19.2       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 3.70  | U         | 3.70     | 19.2       | ug/kg             |
| Total PCBs        | Total PCBs           | 150   |           | 3.60     | 19.2       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 21.5  |           | 32 - 144 | 108%       | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 24.1  |           | 32 - 175 | 121%       | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-18B               |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-09               |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 87.1       | Decanted: |
| Sample Wt/Vol:     | 30.04                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PO113440.D        | 1         | 09/03/25 09:10 | 09/03/25 18:02 | PB169519      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 4.50  | U         | 4.50     | 19.5       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 4.60  | U         | 4.60     | 19.5       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 4.30  | U         | 4.30     | 19.5       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 4.60  | U         | 4.60     | 19.5       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 6.80  | U         | 6.80     | 19.5       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 93.8  |           | 3.70     | 19.5       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 5.80  | U         | 5.80     | 19.5       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 4.10  | U         | 4.10     | 19.5       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 3.70  | U         | 3.70     | 19.5       | ug/kg             |
| Total PCBs        | Total PCBs           | 93.8  |           | 3.70     | 19.5       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 20.9  |           | 32 - 144 | 105%       | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 23.4  |           | 32 - 175 | 117%       | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-19A               |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-11               |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 86.6       | Decanted: |
| Sample Wt/Vol:     | 30.09                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PP074546.D        | 1         | 08/21/25 09:35 | 08/21/25 16:28 | PB169337      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 4.50  | U         | 4.50     | 19.6       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 4.60  | U         | 4.60     | 19.6       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 4.30  | U         | 4.30     | 19.6       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 4.60  | U         | 4.60     | 19.6       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 6.80  | U         | 6.80     | 19.6       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 2000  | E         | 3.70     | 19.6       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 5.80  | U         | 5.80     | 19.6       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 4.10  | U         | 4.10     | 19.6       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 3.70  | U         | 3.70     | 19.6       | ug/kg             |
| Total PCBs        | Total PCBs           | 2000  |           | 3.70     | 19.6       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 17.0  |           | 32 - 144 | 85%        | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 18.6  |           | 32 - 175 | 93%        | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-19ADL             |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-11DL             |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 86.6       | Decanted: |
| Sample Wt/Vol:     | 30.09                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PP074568.D        | 10        | 08/21/25 09:35 | 08/22/25 10:28 | PB169337      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 45.5  | UD        | 45.5     | 196        | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 46.4  | UD        | 46.4     | 196        | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 42.8  | UD        | 42.8     | 196        | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 46.2  | UD        | 46.2     | 196        | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 68.2  | UD        | 68.2     | 196        | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 2000  | D         | 37.0     | 196        | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 57.8  | UD        | 57.8     | 196        | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 41.4  | UD        | 41.4     | 196        | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 37.2  | UD        | 37.2     | 196        | ug/kg             |
| Total PCBs        | Total PCBs           | 2000  | D         | 37.0     | 196        | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 20.4  |           | 32 - 144 | 102%       | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 18.3  |           | 32 - 175 | 92%        | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-19B               |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-12               |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 90.9       | Decanted: |
| Sample Wt/Vol:     | 30.07                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PO113441.D        | 1         | 09/03/25 09:10 | 09/03/25 18:19 | PB169519      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 4.30  | U         | 4.30     | 18.7       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 4.40  | U         | 4.40     | 18.7       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 4.10  | U         | 4.10     | 18.7       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 4.40  | U         | 4.40     | 18.7       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 6.50  | U         | 6.50     | 18.7       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 88.0  |           | 3.50     | 18.7       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 5.50  | U         | 5.50     | 18.7       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 4.00  | U         | 4.00     | 18.7       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 3.50  | U         | 3.50     | 18.7       | ug/kg             |
| Total PCBs        | Total PCBs           | 88.0  |           | 3.50     | 18.7       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 19.0  |           | 32 - 144 | 95%        | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 34.4  |           | 32 - 175 | 172%       | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-20A               |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-14               |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 92.5       | Decanted: |
| Sample Wt/Vol:     | 30.07                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PP074547.D        | 1         | 08/21/25 09:35 | 08/21/25 16:45 | PB169337      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 4.30  | U         | 4.30     | 18.3       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 4.30  | U         | 4.30     | 18.3       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 4.00  | U         | 4.00     | 18.3       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 4.30  | U         | 4.30     | 18.3       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 6.40  | U         | 6.40     | 18.3       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 374   | E         | 3.50     | 18.3       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 5.40  | U         | 5.40     | 18.3       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 3.90  | U         | 3.90     | 18.3       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 3.50  | U         | 3.50     | 18.3       | ug/kg             |
| Total PCBs        | Total PCBs           | 374   |           | 3.50     | 18.3       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 21.5  |           | 32 - 144 | 107%       | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 33.7  |           | 32 - 175 | 169%       | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-20ADL             |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-14DL             |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 92.5       | Decanted: |
| Sample Wt/Vol:     | 30.07                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PP074569.D        | 2         | 08/21/25 09:35 | 08/22/25 10:44 | PB169337      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 8.50  | UD        | 8.50     | 36.7       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 8.70  | UD        | 8.70     | 36.7       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 8.00  | UD        | 8.00     | 36.7       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 8.70  | UD        | 8.70     | 36.7       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 12.8  | UD        | 12.8     | 36.7       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 372   | D         | 6.90     | 36.7       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 10.8  | UD        | 10.8     | 36.7       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 7.80  | UD        | 7.80     | 36.7       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 7.00  | UD        | 7.00     | 36.7       | ug/kg             |
| Total PCBs        | Total PCBs           | 372   | D         | 6.90     | 36.7       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 21.5  |           | 32 - 144 | 108%       | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 29.9  |           | 32 - 175 | 150%       | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-20B               |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-15               |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 91.1       | Decanted: |
| Sample Wt/Vol:     | 30.06                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PP074831.D        | 1         | 09/03/25 09:10 | 09/03/25 20:43 | PB169519      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 4.30  | U         | 4.30     | 18.6       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 4.40  | U         | 4.40     | 18.6       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 4.10  | U         | 4.10     | 18.6       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 4.40  | U         | 4.40     | 18.6       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 6.50  | U         | 6.50     | 18.6       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 3.50  | U         | 3.50     | 18.6       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 5.50  | U         | 5.50     | 18.6       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 3.90  | U         | 3.90     | 18.6       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 3.50  | U         | 3.50     | 18.6       | ug/kg             |
| Total PCBs        | Total PCBs           | 6.50  | U         | 6.50     | 18.6       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 19.5  |           | 32 - 144 | 97%        | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 23.2  |           | 32 - 175 | 116%       | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-21A               |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-17               |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 93.4       | Decanted: |
| Sample Wt/Vol:     | 30.08                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PP074548.D        | 1         | 08/21/25 09:35 | 08/21/25 17:01 | PB169337      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 4.20  | U         | 4.20     | 18.2       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 4.30  | U         | 4.30     | 18.2       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 4.00  | U         | 4.00     | 18.2       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 4.30  | U         | 4.30     | 18.2       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 6.30  | U         | 6.30     | 18.2       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 936   | E         | 3.40     | 18.2       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 5.40  | U         | 5.40     | 18.2       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 3.80  | U         | 3.80     | 18.2       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 3.40  | U         | 3.40     | 18.2       | ug/kg             |
| Total PCBs        | Total PCBs           | 936   |           | 3.40     | 18.2       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 22.0  |           | 32 - 144 | 110%       | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 32.2  |           | 32 - 175 | 161%       | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-21ADL             |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-17DL             |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 93.4       | Decanted: |
| Sample Wt/Vol:     | 30.08                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PP074570.D        | 5         | 08/21/25 09:35 | 08/22/25 11:00 | PB169337      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 21.1  | UD        | 21.1     | 90.8       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 21.5  | UD        | 21.5     | 90.8       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 19.9  | UD        | 19.9     | 90.8       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 21.4  | UD        | 21.4     | 90.8       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 31.6  | UD        | 31.6     | 90.8       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 821   | D         | 17.1     | 90.8       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 26.8  | UD        | 26.8     | 90.8       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 19.2  | UD        | 19.2     | 90.8       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 17.2  | UD        | 17.2     | 90.8       | ug/kg             |
| Total PCBs        | Total PCBs           | 821   | D         | 17.1     | 90.8       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 22.2  |           | 32 - 144 | 111%       | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 26.6  |           | 32 - 175 | 133%       | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-21B               |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-18               |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 91.5       | Decanted: |
| Sample Wt/Vol:     | 30.01                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PP074817.D        | 1         | 09/03/25 09:10 | 09/03/25 15:33 | PB169519      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 4.30  | U         | 4.30     | 18.6       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 4.40  | U         | 4.40     | 18.6       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 4.10  | U         | 4.10     | 18.6       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 4.40  | U         | 4.40     | 18.6       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 6.50  | U         | 6.50     | 18.6       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 3.50  | U         | 3.50     | 18.6       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 5.50  | U         | 5.50     | 18.6       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 3.90  | U         | 3.90     | 18.6       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 3.50  | U         | 3.50     | 18.6       | ug/kg             |
| Total PCBs        | Total PCBs           | 6.50  | U         | 6.50     | 18.6       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 18.5  |           | 32 - 144 | 93%        | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 16.7  |           | 32 - 175 | 84%        | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-22A               |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-20               |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 87.1       | Decanted: |
| Sample Wt/Vol:     | 30.05                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PP074549.D        | 1         | 08/21/25 09:35 | 08/21/25 17:17 | PB169337      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 4.50  | U         | 4.50     | 19.5       | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 4.60  | U         | 4.60     | 19.5       | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 4.30  | U         | 4.30     | 19.5       | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 4.60  | U         | 4.60     | 19.5       | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 6.80  | U         | 6.80     | 19.5       | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 1500  | E         | 3.70     | 19.5       | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 5.80  | U         | 5.80     | 19.5       | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 4.10  | U         | 4.10     | 19.5       | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 3.70  | U         | 3.70     | 19.5       | ug/kg             |
| Total PCBs        | Total PCBs           | 1500  |           | 3.70     | 19.5       | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 20.4  |           | 32 - 144 | 102%       | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 21.6  |           | 32 - 175 | 108%       | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## Report of Analysis

|                    |                        |          |                    |            |           |
|--------------------|------------------------|----------|--------------------|------------|-----------|
| Client:            | ATC Group Services LLC |          | Date Collected:    | 08/19/25   |           |
| Project:           | SCA-K262               |          | Date Received:     | 08/20/25   |           |
| Client Sample ID:  | K262-22ADL             |          | SDG No.:           | Q2922      |           |
| Lab Sample ID:     | Q2922-20DL             |          | Matrix:            | SOIL       |           |
| Analytical Method: | 8082A                  |          | % Solid:           | 87.1       | Decanted: |
| Sample Wt/Vol:     | 30.05                  | Units: g | Final Vol:         | 10000      | uL        |
| Soil Aliquot Vol:  |                        | uL       | Test:              | PCB Group1 |           |
| Extraction Type:   |                        |          | Injection Volume : |            |           |
| GPC Factor :       | 1.0                    | PH :     |                    |            |           |
| Prep Method :      | SW3541B                |          |                    |            |           |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| PP074571.D        | 10        | 08/21/25 09:35 | 08/22/25 11:16 | PB169337      |

| CAS Number        | Parameter            | Conc. | Qualifier | MDL      | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| <b>TARGETS</b>    |                      |       |           |          |            |                   |
| 12674-11-2        | Aroclor-1016         | 45.3  | UD        | 45.3     | 195        | ug/kg             |
| 11104-28-2        | Aroclor-1221         | 46.2  | UD        | 46.2     | 195        | ug/kg             |
| 11141-16-5        | Aroclor-1232         | 42.6  | UD        | 42.6     | 195        | ug/kg             |
| 53469-21-9        | Aroclor-1242         | 46.0  | UD        | 46.0     | 195        | ug/kg             |
| 12672-29-6        | Aroclor-1248         | 67.9  | UD        | 67.9     | 195        | ug/kg             |
| 11097-69-1        | Aroclor-1254         | 1300  | D         | 36.8     | 195        | ug/kg             |
| 37324-23-5        | Aroclor-1262         | 57.5  | UD        | 57.5     | 195        | ug/kg             |
| 11100-14-4        | Aroclor-1268         | 41.3  | UD        | 41.3     | 195        | ug/kg             |
| 11096-82-5        | Aroclor-1260         | 37.0  | UD        | 37.0     | 195        | ug/kg             |
| Total PCBs        | Total PCBs           | 1300  | D         | 36.8     | 195        | ug/kg             |
| <b>SURROGATES</b> |                      |       |           |          |            |                   |
| 877-09-8          | Tetrachloro-m-xylene | 22.7  |           | 32 - 144 | 114%       | SPK: 20           |
| 2051-24-3         | Decachlorobiphenyl   | 19.3  |           | 32 - 175 | 97%        | SPK: 20           |

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates &gt;25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

## LAB CHRONICLE

|                 |                        |                   |                      |
|-----------------|------------------------|-------------------|----------------------|
| <b>OrderID:</b> | Q2922                  | <b>OrderDate:</b> | 8/20/2025 3:10:00 PM |
| <b>Client:</b>  | ATC Group Services LLC | <b>Project:</b>   | SCA-K262             |
| <b>Contact:</b> | Denise Cosenza         | <b>Location:</b>  | J22                  |

| LabID             | ClientID          | Matrix      | Test       | Method | Sample Date     | Prep Date | Anal Date | Received        |
|-------------------|-------------------|-------------|------------|--------|-----------------|-----------|-----------|-----------------|
| <b>Q2922-02</b>   | <b>K262-16A</b>   | <b>SOIL</b> | PCB Group1 | 8082A  | <b>08/19/25</b> | 08/21/25  | 08/21/25  | <b>08/20/25</b> |
| <b>Q2922-03</b>   | <b>K262-16B</b>   | <b>SOIL</b> | PCB Group1 | 8082A  | <b>08/19/25</b> | 09/03/25  | 09/03/25  | <b>08/20/25</b> |
| <b>Q2922-05</b>   | <b>K262-17A</b>   | <b>SOIL</b> | PCB Group1 | 8082A  | <b>08/19/25</b> | 08/21/25  | 08/21/25  | <b>08/20/25</b> |
| <b>Q2922-06</b>   | <b>K262-17B</b>   | <b>SOIL</b> | PCB Group1 | 8082A  | <b>08/19/25</b> | 09/03/25  | 09/03/25  | <b>08/20/25</b> |
| <b>Q2922-08</b>   | <b>K262-18A</b>   | <b>SOIL</b> | PCB Group1 | 8082A  | <b>08/19/25</b> | 08/21/25  | 08/21/25  | <b>08/20/25</b> |
| <b>Q2922-09</b>   | <b>K262-18B</b>   | <b>SOIL</b> | PCB Group1 | 8082A  | <b>08/19/25</b> | 09/03/25  | 09/03/25  | <b>08/20/25</b> |
| <b>Q2922-11</b>   | <b>K262-19A</b>   | <b>SOIL</b> | PCB Group1 | 8082A  | <b>08/19/25</b> | 08/21/25  | 08/21/25  | <b>08/20/25</b> |
| <b>Q2922-11DL</b> | <b>K262-19ADL</b> | <b>SOIL</b> | PCB Group1 | 8082A  | <b>08/19/25</b> | 08/21/25  | 08/22/25  | <b>08/20/25</b> |
| <b>Q2922-12</b>   | <b>K262-19B</b>   | <b>SOIL</b> | PCB Group1 | 8082A  | <b>08/19/25</b> | 09/03/25  | 09/03/25  | <b>08/20/25</b> |
| <b>Q2922-14</b>   | <b>K262-20A</b>   | <b>SOIL</b> | PCB Group1 | 8082A  | <b>08/19/25</b> | 08/21/25  | 08/21/25  | <b>08/20/25</b> |
| <b>Q2922-14DL</b> | <b>K262-20ADL</b> | <b>SOIL</b> | PCB Group1 | 8082A  | <b>08/19/25</b> | 08/21/25  | 08/22/25  | <b>08/20/25</b> |
| <b>Q2922-15</b>   | <b>K262-20B</b>   | <b>SOIL</b> |            |        | <b>08/19/25</b> |           |           | <b>08/20/25</b> |

### LAB CHRONICLE

|                   |                   |             |            |       |                 |          |          |                 |
|-------------------|-------------------|-------------|------------|-------|-----------------|----------|----------|-----------------|
|                   |                   |             | PCB Group1 | 8082A |                 | 09/03/25 | 09/03/25 |                 |
| <b>Q2922-17</b>   | <b>K262-21A</b>   | <b>SOIL</b> |            |       | <b>08/19/25</b> |          |          | <b>08/20/25</b> |
|                   |                   |             | PCB Group1 | 8082A |                 | 08/21/25 | 08/21/25 |                 |
| <b>Q2922-17DL</b> | <b>K262-21ADL</b> | <b>SOIL</b> |            |       | <b>08/19/25</b> |          |          | <b>08/20/25</b> |
|                   |                   |             | PCB Group1 | 8082A |                 | 08/21/25 | 08/22/25 |                 |
| <b>Q2922-18</b>   | <b>K262-21B</b>   | <b>SOIL</b> |            |       | <b>08/19/25</b> |          |          | <b>08/20/25</b> |
|                   |                   |             | PCB Group1 | 8082A |                 | 09/03/25 | 09/03/25 |                 |
| <b>Q2922-20</b>   | <b>K262-22A</b>   | <b>SOIL</b> |            |       | <b>08/19/25</b> |          |          | <b>08/20/25</b> |
|                   |                   |             | PCB Group1 | 8082A |                 | 08/21/25 | 08/21/25 |                 |
| <b>Q2922-20DL</b> | <b>K262-22ADL</b> | <b>SOIL</b> |            |       | <b>08/19/25</b> |          |          | <b>08/20/25</b> |
|                   |                   |             | PCB Group1 | 8082A |                 | 08/21/25 | 08/22/25 |                 |



# SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Atlas  
ADDRESS: 104 E 25th St  
CITY: NY STATE: NY ZIP: 10010  
ATTENTION: d. cozenza  
PHONE: 2123538280 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: SCA-K262  
PROJECT NO.: 2035CA004 LOCATION: Brooklyn  
PROJECT MANAGER: d. cozenza  
e-mail: dcozenza@oretha.com  
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: PO#: SA  
ADDRESS: SA  
CITY: STATE: ZIP:  
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) \_\_\_\_\_ DAYS\*  
HARDCOPY (DATA PACKAGE): 5 day DAYS\*  
EDD: Standard DAYS\*  
\*TO BE APPROVED BY CHEMTECH  
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)  
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP  
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B  
+ Raw Data ☐ Other \_\_\_\_\_  
☐ EDD FORMAT

| 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 |
|---|---|---|---|---|---|---|---|---|
|   |   |   |   |   |   |   |   |   |

| ALLIANCE<br>SAMPLE<br>ID | PROJECT<br>SAMPLE IDENTIFICATION | SAMPLE<br>MATRIX | SAMPLE<br>TYPE |      | SAMPLE<br>COLLECTION |      | # OF BOTTLES | PRESERVATIVES |   |   |   |   |   |   |   |   | COMMENTS |
|--------------------------|----------------------------------|------------------|----------------|------|----------------------|------|--------------|---------------|---|---|---|---|---|---|---|---|----------|
|                          |                                  |                  | COMP           | GRAB | DATE                 | TIME |              | 1             | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 |          |
| 1.                       | K262-15C                         | S                | X              |      | 8/19/25              | 1030 | 1            | X             |   |   |   |   |   |   |   |   | HOLD     |
| 2.                       | K262-16A                         |                  | X              |      | 11                   | 1030 | 1            | X             |   |   |   |   |   |   |   |   | HOLD     |
| 3.                       | K262-16B                         |                  | X              |      | 11                   | 1034 | 1            | X             |   |   |   |   |   |   |   |   | HOLD     |
| 4.                       | K262-16C                         |                  | X              |      | 11                   | 1036 | 1            | X             |   |   |   |   |   |   |   |   | HOLD     |
| 5.                       | K262-17A                         |                  | X              |      | 11                   | 1040 | 1            | X             |   |   |   |   |   |   |   |   | HOLD     |
| 6.                       | K262-178B                        |                  | X              |      | 11                   | 1042 | 1            | X             |   |   |   |   |   |   |   |   | HOLD     |
| 7.                       | K262-17C                         |                  | X              |      | 11                   | 1044 | 1            | X             |   |   |   |   |   |   |   |   | HOLD     |
| 8.                       | K262-18A                         |                  | X              |      | 11                   | 1048 | 1            | X             |   |   |   |   |   |   |   |   | HOLD     |
| 9.                       | K262-18B                         |                  | X              |      | 11                   | 1050 | 1            | X             |   |   |   |   |   |   |   |   | HOLD     |
| 10.                      | K262-18C                         |                  | X              |      | 11                   | 1052 | 1            | X             |   |   |   |   |   |   |   |   | HOLD     |

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

|                          |                |                         |                                                                                                                                                             |
|--------------------------|----------------|-------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------|
| RELINQUISHED BY SAMPLER: | DATE/TIME:     | RECEIVED BY:            | Conditions of bottles or coolers at receipt: <input type="checkbox"/> COMPLIANT <input type="checkbox"/> NON COMPLIANT <input type="checkbox"/> COOLER TEMP |
| 1. <u>[Signature]</u>    | <u>8/19/25</u> | <u>[Signature]</u> 1342 | <u>3.10</u> °C                                                                                                                                              |
| RELINQUISHED BY SAMPLER: | DATE/TIME:     | RECEIVED BY:            | Comments:                                                                                                                                                   |
| 2. <u>[Signature]</u>    | <u>8-20-25</u> | <u>[Signature]</u>      |                                                                                                                                                             |
| RELINQUISHED BY SAMPLER: | DATE/TIME:     | RECEIVED BY:            |                                                                                                                                                             |
| 3. <u>[Signature]</u>    | <u>8-20-25</u> | <u>[Signature]</u>      |                                                                                                                                                             |

Page 5 of 8

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete  
☐ YES ☐ NO



CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

REPORT TO BE SENT TO:

COMPANY: **Atlas**  
ADDRESS: **104 E 25th St**  
CITY: **Ny** STATE: **Ny** ZIP: **10010**  
ATTENTION: **d. cosenza**  
PHONE: **2123538280** FAX:

PROJECT NAME: **SCA-K262**  
PROJECT NO.: **2025SCA024** LOCATION: **Backlin**  
PROJECT MANAGER: **d. cosenza**  
e-mail: **claire.cosenza@atlas.com**  
PHONE: FAX:

BILL TO: PO#: ADDRESS: CITY STATE: ZIP: ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

DATA DELIVERABLE INFORMATION

FAX (RUSH) DAYS\*  
HARDCOPY (DATA PACKAGE): **Standard** DAYS\*  
EDD: **5 DAY** DAYS\*  
\*TO BE APPROVED BY CHEMTECH  
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)  
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP  
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B  
+ Raw Data ☐ Other  
☐ EDD FORMAT

**Total PCBs**

|   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|
| 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 |
|---|---|---|---|---|---|---|---|---|

PRESERVATIVES

COMMENTS

| ALLIANCE<br>SAMPLE<br>ID | PROJECT<br>SAMPLE IDENTIFICATION | SAMPLE<br>MATRIX | SAMPLE<br>TYPE |      | SAMPLE<br>COLLECTION |       | # OF BOTTLES |   |   |   |   |   |   |   |   |   | ← Specify Preservatives<br>A-HCl D-NaOH<br>B-HNO3 E-ICE<br>C-H2SO4 F-OTHER |
|--------------------------|----------------------------------|------------------|----------------|------|----------------------|-------|--------------|---|---|---|---|---|---|---|---|---|----------------------------------------------------------------------------|
|                          |                                  |                  | COMP           | GRAB | DATE                 | TIME  |              | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 |                                                                            |
| 1.                       | K262-19A                         | S                | X              | X    | 8/19/25              | 11:00 | 1            | X |   |   |   |   |   |   |   |   |                                                                            |
| 2.                       | K262-19B                         |                  | X              | X    | 11                   | 11:02 | 1            | X |   |   |   |   |   |   |   |   | HOLD                                                                       |
| 3.                       | K262-19C                         |                  | X              | X    | 11                   | 11:05 | 1            | X |   |   |   |   |   |   |   |   | HOLD                                                                       |
| 4.                       | K262-20A                         |                  | X              | X    | 11                   | 11:10 | 1            | X |   |   |   |   |   |   |   |   |                                                                            |
| 5.                       | K262-20B                         |                  | X              | X    | 11                   | 11:11 | 1            | X |   |   |   |   |   |   |   |   | HOLD                                                                       |
| 6.                       | K262-20C                         |                  | X              | X    | 11                   | 11:13 | 1            | X |   |   |   |   |   |   |   |   | HOLD                                                                       |
| 7.                       | K262-21A                         |                  | X              | X    | 11                   | 11:15 | 1            | X |   |   |   |   |   |   |   |   |                                                                            |
| 8.                       | K262-21B                         |                  | X              | X    | 11                   | 11:17 | 1            | X |   |   |   |   |   |   |   |   | HOLD                                                                       |
| 9.                       | K262-21C                         |                  | X              | X    | 11                   | 11:19 | 1            | X |   |   |   |   |   |   |   |   | HOLD                                                                       |
| 10.                      | K262-22A                         |                  | X              | X    | 11                   | 11:22 | 1            | X |   |   |   |   |   |   |   |   |                                                                            |

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

|                          |                           |              |                                                                                                                                                                           |
|--------------------------|---------------------------|--------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| RELINQUISHED BY SAMPLER: | DATE/TIME: <b>8/19/25</b> | RECEIVED BY: | Conditions of bottles or coolers at receipt: <input type="checkbox"/> COMPLIANT <input type="checkbox"/> NON COMPLIANT <input type="checkbox"/> COOLER TEMP <b>3.1</b> °C |
| 1.                       |                           | 1.           | Comments:                                                                                                                                                                 |
| RELINQUISHED BY SAMPLER: | DATE/TIME: <b>8</b>       | RECEIVED BY: |                                                                                                                                                                           |
| 2.                       |                           | 2.           |                                                                                                                                                                           |
| RELINQUISHED BY SAMPLER: | DATE/TIME: <b>1729</b>    | RECEIVED BY: |                                                                                                                                                                           |
| 3.                       | <b>8-20-25</b>            | 3.           |                                                                                                                                                                           |

Page **6** of **8**

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete  
☐ YES ☐ NO



---

**From:** Denise Cosenza <denise.cosenza@oneatlas.com>  
**Sent:** Friday, September 05, 2025 9:12 AM  
**To:** Yazmeen Gomez  
**Cc:** Peter De Garay  
**Subject:** RE: [EXTERNAL] RE: SCA - PS K262 PCB Soil Sampling Glassware

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good morning Yazmeen,

We will not be analyzing the remaining samples for this project.

Any questions, please reach out.

Denise



**Denise Cosenza**

Project Manager, Environmental Division

O: 212.284.0613 | C: 718.490.0614

---

**From:** Denise Cosenza  
**Sent:** Thursday, August 28, 2025 12:48 PM  
**To:** Yazmeen Gomez <Yazmeen.Gomez@alliancetg.com>  
**Cc:** Peter De Garay <Peter.DeGaray@oneatlas.com>  
**Subject:** RE: [EXTERNAL] RE: SCA - PS K262 PCB Soil Sampling Glassware

Hi Yazmeen,

Is it safe to say that we will not get results during the hold time for remaining samples?

If not, let's extract and hold 16C, 17C, 18C, 19C, 20C, 21C, 22C, 23C, and 24C. Activation is dependent on the analysis results and I will let you know once we have those.

None of the other samples on hold will have to be extracted.

Thank you,  
Denise


**Denise Cosenza**

Project Manager, Environmental Division

O: 212.284.0613 | C: 718.490.0614

**From:** Yazmeen Gomez <[Yazmeen.Gomez@alliancetg.com](mailto:Yazmeen.Gomez@alliancetg.com)>  
**Sent:** Thursday, August 28, 2025 12:42 PM  
**To:** Denise Cosenza <[denise.cosenza@oneatlas.com](mailto:denise.cosenza@oneatlas.com)>  
**Cc:** Peter De Garay <[Peter.DeGaray@oneatlas.com](mailto:Peter.DeGaray@oneatlas.com)>  
**Subject:** RE: [EXTERNAL] RE: SCA - PS K262 PCB Soil Sampling Glassware

Hi Denise,

Confirming these samples have been activated as requested.

**Best Regards,**



**Yazmeen** Gomez  
**Sr. Project Manager**  
**An Alliance Technical Group Company**  
**Main:** 908-789-8900  
**Direct:** 908-728-3147  
**Address:** 284 Sheffield St, Ste 1, Mountainside, NJ 07092  
[www.alliancetg.com](http://www.alliancetg.com)

**From:** Denise Cosenza <[denise.cosenza@oneatlas.com](mailto:denise.cosenza@oneatlas.com)>  
**Sent:** Thursday, August 28, 2025 10:55 AM  
**To:** Yazmeen Gomez <[Yazmeen.Gomez@alliancetg.com](mailto:Yazmeen.Gomez@alliancetg.com)>  
**Cc:** Peter De Garay <[Peter.DeGaray@oneatlas.com](mailto:Peter.DeGaray@oneatlas.com)>  
**Subject:** RE: [EXTERNAL] RE: SCA - PS K262 PCB Soil Sampling Glassware

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Hi Yazmeen,

Based on the results received, I would like to activate the following samples for total pcb analysis for this project:

16B, 17B, 18B, 19B, 20B, 21B, 22B, 23B, 24B and DUP 6.

Thank you,  
 Denise


**Denise Cosenza**

Project Manager, Environmental Division

O: 212.284.0613 | C: 718.490.0614

---

**From:** Yazmeen Gomez <[Yazmeen.Gomez@alliancetg.com](mailto:Yazmeen.Gomez@alliancetg.com)>  
**Sent:** Tuesday, August 19, 2025 1:07 PM  
**To:** Denise Cosenza <[denise.cosenza@oneatlas.com](mailto:denise.cosenza@oneatlas.com)>  
**Cc:** Peter De Garay <[Peter.DeGaray@oneatlas.com](mailto:Peter.DeGaray@oneatlas.com)>  
**Subject:** RE: [EXTERNAL] RE: SCA - PS K262 PCB Soil Sampling Glassware

Hi Denise,

Pick up for tomorrow confirmed.

**Best Regards,**



**Yazmeen** Gomez

**Sr. Project Manager**

**An Alliance Technical Group Company**

**Main:** 908-789-8900

**Direct:** 908-728-3147

**Address:** 284 Sheffield St, Ste 1, Mountainside, NJ 07092

[www.alliancetg.com](http://www.alliancetg.com)     

---

**From:** Denise Cosenza <[denise.cosenza@oneatlas.com](mailto:denise.cosenza@oneatlas.com)>  
**Sent:** Tuesday, August 19, 2025 1:02 PM  
**To:** Yazmeen Gomez <[Yazmeen.Gomez@alliancetg.com](mailto:Yazmeen.Gomez@alliancetg.com)>  
**Cc:** Peter De Garay <[Peter.DeGaray@oneatlas.com](mailto:Peter.DeGaray@oneatlas.com)>  
**Subject:** RE: [EXTERNAL] RE: SCA - PS K262 PCB Soil Sampling Glassware

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Hi Yazmeen,

These samples were collected today and will be available for pickup tomorrow from our Bethpage office.

**Peter deGaray**

**Atlas Technical Consultants**

999 S Oyster Bay Road, Suite 114

Bethpage, NY, 11746

C: [631.901.7390](tel:631.901.7390)

[peter.degaray@oneatlas.com](mailto:peter.degaray@oneatlas.com)

Thank you,  
Denise

Denise Cosenza  
Project Manager  
ATLAS  
C. 718.490.0614  
Sent from my Verizon, Samsung Galaxy smartphone

----- Original message -----

From: Yazmeen Gomez <[Yazmeen.Gomez@alliancetg.com](mailto:Yazmeen.Gomez@alliancetg.com)>  
Date: 8/11/25 2:28 PM (GMT-05:00)  
To: Denise Cosenza <[denise.cosenza@oneatlas.com](mailto:denise.cosenza@oneatlas.com)>  
Cc: Peter De Garay <[Peter.DeGaray@oneatlas.com](mailto:Peter.DeGaray@oneatlas.com)>  
Subject: [EXTERNAL] RE: SCA - PS K262 PCB Soil Sampling Glassware

**[External Email]** This email originated from outside of the Atlas mail system. Please use caution when opening attachments.

Good afternoon Denise,

Bottle order delivery confirmed for tomorrow.

Please let me know if you need anything else, have a good week!

**Best Regards,**



**Yazmeen** Gomez  
**Sr. Project Manager**  
**An Alliance Technical Group Company**  
**Main:** 908-789-8900  
**Direct:** 908-728-3147  
**Address:** 284 Sheffield St, Ste 1, Mountainside, NJ 07092  
[www.alliancetg.com](http://www.alliancetg.com)   

---

**From:** Denise Cosenza <[denise.cosenza@oneatlas.com](mailto:denise.cosenza@oneatlas.com)>  
**Sent:** Monday, August 11, 2025 2:16 PM  
**To:** Yazmeen Gomez <[Yazmeen.Gomez@alliancetg.com](mailto:Yazmeen.Gomez@alliancetg.com)>  
**Cc:** Peter De Garay <[Peter.DeGaray@oneatlas.com](mailto:Peter.DeGaray@oneatlas.com)>  
**Subject:** SCA - PS K262 PCB Soil Sampling Glassware

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon Yazmeen,

I'd like to place the following order for glassware:

Project Name: SCA K262

Project No. 2025sca024

76 soil samples – Total PCB analysis

Deliver to Peter at the Bethpage Office:

**Peter deGaray**

**Atlas Technical Consultants**

999 S Oyster Bay Road, Suite 114

Bethpage, NY, 11746

C: 631.901.7390

[peter.degaray@oneatlas.com](mailto:peter.degaray@oneatlas.com)

Please deliver this week.

Proposed Sampling Date: Next Week

Any questions Please let me know,

Thank you!

Denise

**Denise Cosenza**

Project Manager, Environmental Division



**Atlas Technical Consultants**

104 East 25<sup>th</sup> Street, 8<sup>th</sup> Floor

New York, NY 10010

O: 212.284.0613 | C: 718.490.0614

[oneatlas.com](http://oneatlas.com)

in f X



This message and any attachments are intended only for the use of the addressee and may contain information that is privileged and/or confidential. If the reader of the message is not the intended recipient or an authorized representative of the intended recipient, you are hereby notified that any use and/or dissemination of any of this communication is strictly prohibited. If you have received this communication in error, notify the sender immediately by return email and delete the message and any attachments from your system.

This message and any attachments are intended only for the use of the addressee and may contain information that is privileged and/or confidential. If the reader of the message is not the intended recipient or an authorized representative of the intended recipient, you are hereby notified that any use and/or dissemination of any of this communication is strictly prohibited. If you have received this communication in error, notify the sender immediately by return email and delete the message and any attachments from your system.

This message and any attachments are intended only for the use of the addressee and may contain information that is privileged and/or confidential. If the reader of the message is not the intended recipient or an authorized representative of the intended recipient, you are hereby notified that any use and/or dissemination of any of this communication is strictly prohibited. If you have received this communication in error, notify the sender immediately by return email and delete the message and any attachments from your system.

This message and any attachments are intended only for the use of the addressee and may contain information that is privileged and/or confidential. If the reader of the message is not the intended recipient or an authorized representative of the intended recipient, you are hereby notified that any use and/or dissemination of any of this communication is strictly prohibited. If you have received this communication in error, notify the sender immediately by return email and delete the message and any attachments from your system.

### Laboratory Certification

| Certified By    | License No.      |
|-----------------|------------------|
|                 |                  |
| Connecticut     | PH-0830          |
|                 |                  |
| DOD ELAP (ANAB) | L2219            |
|                 |                  |
| Maine           | 2024021          |
|                 |                  |
| Maryland        | 296              |
|                 |                  |
| New Hampshire   | 255425           |
|                 |                  |
| New Jersey      | 20012            |
|                 |                  |
| New York        | 11376            |
|                 |                  |
| Pennsylvania    | 68-00548         |
|                 |                  |
| Soil Permit     | 525-24-234-08441 |
|                 |                  |
| Texas           | TX-C25-00189     |
|                 |                  |
| Virginia        | 460312           |