

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2922

Continuing Calib Date: 08/21/2025

Initial Calibration Date(s): 08/19/2025

08/19/2025

Continuing Calib Time: 21:21

Initial Calibration Time(s): 10:02

17:36

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1254-1       | (1) | 6.66       | 6.66      | 6.56      | 6.76  | 0.00       |
| Aroclor-1254-2       | (2) | 6.88       | 6.87      | 6.77      | 6.97  | -0.01      |
| Aroclor-1254-3       | (3) | 7.24       | 7.24      | 7.14      | 7.34  | 0.00       |
| Aroclor-1254-4       | (4) | 7.52       | 7.52      | 7.42      | 7.62  | 0.00       |
| Aroclor-1254-5       | (5) | 7.93       | 7.93      | 7.83      | 8.03  | 0.00       |
| Tetrachloro-m-xylene |     | 4.66       | 4.66      | 4.56      | 4.76  | 0.00       |
| Decachlorobiphenyl   |     | 10.44      | 10.44     | 10.34     | 10.54 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2922

Continuing Calib Date: 08/21/2025

Initial Calibration Date(s): 08/19/2025

08/19/2025

Continuing Calib Time: 21:21

Initial Calibration Time(s): 10:02

17:36

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.68       | 5.68      | 5.58      | 5.78 | 0.00       |
| Aroclor-1254-2 (2)   | 5.83       | 5.83      | 5.73      | 5.93 | 0.00       |
| Aroclor-1254-3 (3)   | 6.23       | 6.23      | 6.13      | 6.33 | 0.00       |
| Aroclor-1254-4 (4)   | 6.46       | 6.46      | 6.36      | 6.56 | 0.00       |
| Aroclor-1254-5 (5)   | 6.88       | 6.88      | 6.78      | 6.98 | 0.00       |
| Tetrachloro-m-xylene | 3.80       | 3.80      | 3.70      | 3.90 | 0.00       |
| Decachlorobiphenyl   | 8.81       | 8.82      | 8.72      | 8.92 | 0.01       |