

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2922
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074541.D **Time Analyzed:** 15:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.811	5.706	5.906	471.190	500.000	-5.8
Aroclor-1016-2	5.832	5.727	5.927	484.200	500.000	-3.2
Aroclor-1016-3	5.895	5.790	5.990	455.310	500.000	-8.9
Aroclor-1016-4	5.992	5.888	6.088	476.030	500.000	-4.8
Aroclor-1016-5	6.285	6.180	6.380	493.590	500.000	-1.3
Aroclor-1260-1	7.402	7.297	7.497	474.380	500.000	-5.1
Aroclor-1260-2	7.654	7.550	7.750	475.830	500.000	-4.8
Aroclor-1260-3	8.012	7.908	8.108	506.310	500.000	1.3
Aroclor-1260-4	8.241	8.136	8.336	490.910	500.000	-1.8
Aroclor-1260-5	8.566	8.462	8.662	477.750	500.000	-4.5
Decachlorobiphenyl	10.437	10.333	10.533	49.420	50.000	-1.2
Tetrachloro-m-xylene	4.659	4.554	4.754	47.070	50.000	-5.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2922
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074541.D **Time Analyzed:** 15:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.901	4.799	4.999	482.690	500.000	-3.5
Aroclor-1016-2	4.958	4.855	5.055	485.710	500.000	-2.9
Aroclor-1016-3	5.079	4.976	5.176	474.360	500.000	-5.1
Aroclor-1016-4	5.119	5.017	5.217	478.130	500.000	-4.4
Aroclor-1016-5	5.333	5.231	5.431	441.000	500.000	-11.8
Aroclor-1260-1	6.364	6.261	6.461	443.160	500.000	-11.4
Aroclor-1260-2	6.550	6.448	6.648	429.440	500.000	-14.1
Aroclor-1260-3	6.702	6.602	6.802	404.020	500.000	-19.2
Aroclor-1260-4	7.174	7.072	7.272	425.360	500.000	-14.9
Aroclor-1260-5	7.413	7.311	7.511	419.550	500.000	-16.1
Decachlorobiphenyl	8.816	8.715	8.915	44.110	50.000	-11.8
Tetrachloro-m-xylene	3.802	3.699	3.899	47.800	50.000	-4.4