

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ATCE02  
**Lab Code:** ACE **SDG NO.:** Q2922  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

**Client Sample No.:** CCAL06 **Date Analyzed:** 08/21/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP074556.D **Time Analyzed:** 19:59

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.811  | 5.706     | 5.906  | 476.060            | 500.000           | -4.8 |
| Aroclor-1016-2       | 5.833  | 5.727     | 5.927  | 493.460            | 500.000           | -1.3 |
| Aroclor-1016-3       | 5.896  | 5.790     | 5.990  | 453.750            | 500.000           | -9.3 |
| Aroclor-1016-4       | 5.993  | 5.888     | 6.088  | 483.490            | 500.000           | -3.3 |
| Aroclor-1016-5       | 6.285  | 6.180     | 6.380  | 501.400            | 500.000           | 0.3  |
| Aroclor-1260-1       | 7.403  | 7.297     | 7.497  | 477.480            | 500.000           | -4.5 |
| Aroclor-1260-2       | 7.656  | 7.550     | 7.750  | 480.720            | 500.000           | -3.9 |
| Aroclor-1260-3       | 8.013  | 7.908     | 8.108  | 514.970            | 500.000           | 3.0  |
| Aroclor-1260-4       | 8.242  | 8.136     | 8.336  | 497.970            | 500.000           | -0.4 |
| Aroclor-1260-5       | 8.568  | 8.462     | 8.662  | 481.020            | 500.000           | -3.8 |
| Decachlorobiphenyl   | 10.439 | 10.333    | 10.533 | 48.860             | 50.000            | -2.3 |
| Tetrachloro-m-xylene | 4.660  | 4.554     | 4.754  | 48.040             | 50.000            | -3.9 |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ATCE02  
**Lab Code:** ACE **SDG NO.:** Q2922  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

**Client Sample No.:** CCAL06 **Date Analyzed:** 08/21/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP074556.D **Time Analyzed:** 19:59

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.902 | 4.799     | 4.999 | 497.550            | 500.000           | -0.5  |
| Aroclor-1016-2       | 4.959 | 4.855     | 5.055 | 505.540            | 500.000           | 1.1   |
| Aroclor-1016-3       | 5.079 | 4.976     | 5.176 | 502.390            | 500.000           | 0.5   |
| Aroclor-1016-4       | 5.120 | 5.017     | 5.217 | 478.010            | 500.000           | -4.4  |
| Aroclor-1016-5       | 5.335 | 5.231     | 5.431 | 449.060            | 500.000           | -10.2 |
| Aroclor-1260-1       | 6.364 | 6.261     | 6.461 | 466.850            | 500.000           | -6.6  |
| Aroclor-1260-2       | 6.550 | 6.448     | 6.648 | 451.180            | 500.000           | -9.8  |
| Aroclor-1260-3       | 6.704 | 6.602     | 6.802 | 418.490            | 500.000           | -16.3 |
| Aroclor-1260-4       | 7.174 | 7.072     | 7.272 | 452.890            | 500.000           | -9.4  |
| Aroclor-1260-5       | 7.413 | 7.311     | 7.511 | 441.700            | 500.000           | -11.7 |
| Decachlorobiphenyl   | 8.816 | 8.715     | 8.915 | 44.520             | 50.000            | -11.0 |
| Tetrachloro-m-xylene | 3.802 | 3.699     | 3.899 | 50.270             | 50.000            | 0.5   |