

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ATCE02  
**Lab Code:** ACE **SDG NO.:** Q2922  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

**Client Sample No.:** CCAL07 **Date Analyzed:** 08/22/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP074562.D **Time Analyzed:** 08:50

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.809  | 5.706     | 5.906  | 485.240            | 500.000           | -3.0 |
| Aroclor-1016-2       | 5.831  | 5.727     | 5.927  | 508.080            | 500.000           | 1.6  |
| Aroclor-1016-3       | 5.894  | 5.790     | 5.990  | 471.790            | 500.000           | -5.6 |
| Aroclor-1016-4       | 5.991  | 5.888     | 6.088  | 492.820            | 500.000           | -1.4 |
| Aroclor-1016-5       | 6.283  | 6.180     | 6.380  | 511.900            | 500.000           | 2.4  |
| Aroclor-1260-1       | 7.401  | 7.297     | 7.497  | 492.530            | 500.000           | -1.5 |
| Aroclor-1260-2       | 7.653  | 7.550     | 7.750  | 489.260            | 500.000           | -2.1 |
| Aroclor-1260-3       | 8.011  | 7.908     | 8.108  | 517.170            | 500.000           | 3.4  |
| Aroclor-1260-4       | 8.239  | 8.136     | 8.336  | 500.020            | 500.000           | 0.0  |
| Aroclor-1260-5       | 8.565  | 8.462     | 8.662  | 492.030            | 500.000           | -1.6 |
| Decachlorobiphenyl   | 10.436 | 10.333    | 10.533 | 50.780             | 50.000            | 1.6  |
| Tetrachloro-m-xylene | 4.657  | 4.554     | 4.754  | 48.970             | 50.000            | -2.1 |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ATCE02  
**Lab Code:** ACE **SDG NO.:** Q2922  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

**Client Sample No.:** CCAL07 **Date Analyzed:** 08/22/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP074562.D **Time Analyzed:** 08:50

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.899 | 4.799     | 4.999 | 496.520            | 500.000           | -0.7  |
| Aroclor-1016-2       | 4.957 | 4.855     | 5.055 | 506.360            | 500.000           | 1.3   |
| Aroclor-1016-3       | 5.077 | 4.976     | 5.176 | 488.410            | 500.000           | -2.3  |
| Aroclor-1016-4       | 5.118 | 5.017     | 5.217 | 500.870            | 500.000           | 0.2   |
| Aroclor-1016-5       | 5.331 | 5.231     | 5.431 | 485.760            | 500.000           | -2.8  |
| Aroclor-1260-1       | 6.362 | 6.261     | 6.461 | 483.650            | 500.000           | -3.3  |
| Aroclor-1260-2       | 6.548 | 6.448     | 6.648 | 483.570            | 500.000           | -3.3  |
| Aroclor-1260-3       | 6.702 | 6.602     | 6.802 | 439.540            | 500.000           | -12.1 |
| Aroclor-1260-4       | 7.172 | 7.072     | 7.272 | 479.380            | 500.000           | -4.1  |
| Aroclor-1260-5       | 7.411 | 7.311     | 7.511 | 464.100            | 500.000           | -7.2  |
| Decachlorobiphenyl   | 8.813 | 8.715     | 8.915 | 48.500             | 50.000            | -3.0  |
| Tetrachloro-m-xylene | 3.800 | 3.699     | 3.899 | 51.200             | 50.000            | 2.4   |