284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2922 **Analytical Method:** 8082A

Client: ATC Group Services LLC **Datafile :** PO113142.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD Qual	Limits Low	Limits High	RPD
PB169337BS (Column 1)	AR1016	166.6	152	ug/kg	91				71	120	
	AR1260	166.6	142	ug/kg	85				65	130	
PB169337BS (Column 2)	AR1016	166.6	145	ug/kg	87				71	120	
	AR1260	166.6	135	ug/kg	81				65	130	