

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2922

Continuing Calib Date: 08/22/2025

Initial Calibration Date(s): 08/18/2025

08/18/2025

Continuing Calib Time: 09:34

Initial Calibration Time(s): 09:53

18:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.75	4.75	4.65	4.85	0.00
Aroclor-1248-2 (2)	4.98	4.98	4.88	5.08	0.00
Aroclor-1248-3 (3)	5.20	5.20	5.10	5.30	0.00
Aroclor-1248-4 (4)	5.55	5.55	5.45	5.65	0.00
Aroclor-1248-5 (5)	5.59	5.59	5.49	5.69	0.00
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl	8.68	8.68	8.58	8.78	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2922

Continuing Calib Date: 08/22/2025

Initial Calibration Date(s): 08/18/2025 08/18/2025

Continuing Calib Time: 09:34

Initial Calibration Time(s): 09:53 18:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.73	4.73	4.63	4.83	0.00
Aroclor-1248-2 (2)	4.96	4.96	4.86	5.06	0.00
Aroclor-1248-3 (3)	5.00	5.01	4.91	5.11	0.01
Aroclor-1248-4 (4)	5.17	5.18	5.08	5.28	0.01
Aroclor-1248-5 (5)	5.56	5.57	5.47	5.67	0.01
Tetrachloro-m-xylene	3.65	3.66	3.56	3.76	0.01
Decachlorobiphenyl	8.62	8.63	8.53	8.73	0.01