

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2922

Continuing Calib Date: 09/03/2025

Initial Calibration Date(s): 08/18/2025 08/18/2025

Continuing Calib Time: 08:56

Initial Calibration Time(s): 09:53 18:24

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.75       | 4.75      | 4.65      | 4.85 | 0.00       |
| Aroclor-1016-2       | (2) | 4.76       | 4.77      | 4.67      | 4.87 | 0.01       |
| Aroclor-1016-3       | (3) | 4.82       | 4.82      | 4.72      | 4.92 | 0.00       |
| Aroclor-1016-4       | (4) | 4.94       | 4.94      | 4.84      | 5.04 | 0.00       |
| Aroclor-1016-5       | (5) | 5.20       | 5.20      | 5.10      | 5.30 | 0.00       |
| Aroclor-1260-1       | (1) | 6.23       | 6.24      | 6.14      | 6.34 | 0.01       |
| Aroclor-1260-2       | (2) | 6.42       | 6.43      | 6.33      | 6.53 | 0.01       |
| Aroclor-1260-3       | (3) | 6.79       | 6.79      | 6.69      | 6.89 | 0.00       |
| Aroclor-1260-4       | (4) | 7.05       | 7.05      | 6.95      | 7.15 | 0.00       |
| Aroclor-1260-5       | (5) | 7.29       | 7.29      | 7.19      | 7.39 | 0.00       |
| Tetrachloro-m-xylene |     | 3.66       | 3.66      | 3.56      | 3.76 | 0.00       |
| Decachlorobiphenyl   |     | 8.68       | 8.68      | 8.58      | 8.78 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2922

Continuing Calib Date: 09/03/2025

Initial Calibration Date(s): 08/18/2025 08/18/2025

Continuing Calib Time: 08:56

Initial Calibration Time(s): 09:53 18:24

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1       | (1) | 4.72       | 4.73      | 4.63                   | 4.83 | 0.01       |
| Aroclor-1016-2       | (2) | 4.74       | 4.75      | 4.65                   | 4.85 | 0.01       |
| Aroclor-1016-3       | (3) | 4.92       | 4.92      | 4.82                   | 5.02 | 0.00       |
| Aroclor-1016-4       | (4) | 4.96       | 4.96      | 4.86                   | 5.06 | 0.00       |
| Aroclor-1016-5       | (5) | 5.17       | 5.18      | 5.08                   | 5.28 | 0.01       |
| Aroclor-1260-1       | (1) | 6.20       | 6.20      | 6.10                   | 6.30 | 0.00       |
| Aroclor-1260-2       | (2) | 6.39       | 6.39      | 6.29                   | 6.49 | 0.00       |
| Aroclor-1260-3       | (3) | 6.54       | 6.54      | 6.44                   | 6.64 | 0.00       |
| Aroclor-1260-4       | (4) | 7.01       | 7.01      | 6.91                   | 7.11 | 0.00       |
| Aroclor-1260-5       | (5) | 7.25       | 7.26      | 7.16                   | 7.36 | 0.01       |
| Tetrachloro-m-xylene |     | 3.65       | 3.66      | 3.56                   | 3.76 | 0.01       |
| Decachlorobiphenyl   |     | 8.62       | 8.63      | 8.53                   | 8.73 | 0.01       |