

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2922

Continuing Calib Date: 08/23/2025

Initial Calibration Date(s): 08/19/2025

08/19/2025

Continuing Calib Time: 01:54

Initial Calibration Time(s): 10:02

17:36

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1248-1 (1)   | 5.81       | 5.81      | 5.71      | 5.91  | 0.00       |
| Aroclor-1248-2 (2)   | 6.08       | 6.08      | 5.98      | 6.18  | 0.00       |
| Aroclor-1248-3 (3)   | 6.28       | 6.28      | 6.18      | 6.38  | 0.00       |
| Aroclor-1248-4 (4)   | 6.68       | 6.68      | 6.58      | 6.78  | 0.00       |
| Aroclor-1248-5 (5)   | 6.72       | 6.72      | 6.62      | 6.82  | 0.00       |
| Tetrachloro-m-xylene | 4.66       | 4.66      | 4.56      | 4.76  | 0.00       |
| Decachlorobiphenyl   | 10.44      | 10.44     | 10.34     | 10.54 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2922

Continuing Calib Date: 08/23/2025

Initial Calibration Date(s): 08/19/2025

08/19/2025

Continuing Calib Time: 01:54

Initial Calibration Time(s): 10:02

17:36

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 4.90       | 4.90      | 4.80      | 5.00 | 0.00       |
| Aroclor-1248-2 (2)   | 5.12       | 5.12      | 5.02      | 5.22 | 0.00       |
| Aroclor-1248-3 (3)   | 5.16       | 5.16      | 5.06      | 5.26 | 0.00       |
| Aroclor-1248-4 (4)   | 5.33       | 5.33      | 5.23      | 5.43 | 0.00       |
| Aroclor-1248-5 (5)   | 5.72       | 5.72      | 5.62      | 5.82 | 0.00       |
| Tetrachloro-m-xylene | 3.80       | 3.80      | 3.70      | 3.90 | 0.00       |
| Decachlorobiphenyl   | 8.81       | 8.82      | 8.72      | 8.92 | 0.01       |