

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2922
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 09/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113428.D **Time Analyzed:** 14:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.745	4.648	4.848	552.320	500.000	10.5
Aroclor-1016-2	4.764	4.666	4.866	549.260	500.000	9.9
Aroclor-1016-3	4.821	4.724	4.924	553.180	500.000	10.6
Aroclor-1016-4	4.940	4.843	5.043	471.990	500.000	-5.6
Aroclor-1016-5	5.196	5.099	5.299	531.930	500.000	6.4
Aroclor-1260-1	6.232	6.136	6.336	483.180	500.000	-3.4
Aroclor-1260-2	6.422	6.326	6.526	460.330	500.000	-7.9
Aroclor-1260-3	6.787	6.691	6.891	439.480	500.000	-12.1
Aroclor-1260-4	7.046	6.951	7.151	420.540	500.000	-15.9
Aroclor-1260-5	7.290	7.194	7.394	405.660	500.000	-18.9
Decachlorobiphenyl	8.676	8.582	8.782	39.650	50.000	-20.7
Tetrachloro-m-xylene	3.661	3.563	3.763	55.650	50.000	11.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2922
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 09/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113428.D **Time Analyzed:** 14:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.629	4.829	517.560	500.000	3.5
Aroclor-1016-2	4.744	4.647	4.847	514.260	500.000	2.9
Aroclor-1016-3	4.919	4.822	5.022	497.670	500.000	-0.5
Aroclor-1016-4	4.961	4.864	5.064	483.010	500.000	-3.4
Aroclor-1016-5	5.173	5.076	5.276	522.440	500.000	4.5
Aroclor-1260-1	6.199	6.104	6.304	434.620	500.000	-13.1
Aroclor-1260-2	6.388	6.292	6.492	423.550	500.000	-15.3
Aroclor-1260-3	6.538	6.443	6.643	398.010	500.000	-20.4
Aroclor-1260-4	7.008	6.912	7.112	403.910	500.000	-19.2
Aroclor-1260-5	7.250	7.155	7.355	404.690	500.000	-19.1
Decachlorobiphenyl	8.621	8.528	8.728	46.000	50.000	-8.0
Tetrachloro-m-xylene	3.652	3.555	3.755	53.080	50.000	6.2