

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ATCE02  
**Lab Code:** ACE **SDG NO.:** Q2922  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

**Client Sample No.:** CCAL12 **Date Analyzed:** 09/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP074819.D **Time Analyzed:** 16:38

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|--------|-----------|--------|--------------------|-------------------|-------|
|                      |        | FROM      | TO     |                    |                   |       |
| Aroclor-1016-1       | 5.813  | 5.706     | 5.906  | 503.390            | 500.000           | 0.7   |
| Aroclor-1016-2       | 5.834  | 5.727     | 5.927  | 519.210            | 500.000           | 3.8   |
| Aroclor-1016-3       | 5.897  | 5.790     | 5.990  | 470.950            | 500.000           | -5.8  |
| Aroclor-1016-4       | 5.994  | 5.888     | 6.088  | 497.160            | 500.000           | -0.6  |
| Aroclor-1016-5       | 6.287  | 6.180     | 6.380  | 515.490            | 500.000           | 3.1   |
| Aroclor-1260-1       | 7.404  | 7.297     | 7.497  | 474.610            | 500.000           | -5.1  |
| Aroclor-1260-2       | 7.657  | 7.550     | 7.750  | 462.930            | 500.000           | -7.4  |
| Aroclor-1260-3       | 8.015  | 7.908     | 8.108  | 487.570            | 500.000           | -2.5  |
| Aroclor-1260-4       | 8.243  | 8.136     | 8.336  | 460.910            | 500.000           | -7.8  |
| Aroclor-1260-5       | 8.569  | 8.462     | 8.662  | 446.490            | 500.000           | -10.7 |
| Decachlorobiphenyl   | 10.440 | 10.333    | 10.533 | 54.250             | 50.000            | 8.5   |
| Tetrachloro-m-xylene | 4.661  | 4.554     | 4.754  | 52.870             | 50.000            | 5.7   |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ATCE02  
**Lab Code:** ACE **SDG NO.:** Q2922  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

**Client Sample No.:** CCAL12 **Date Analyzed:** 09/03/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP074819.D **Time Analyzed:** 16:38

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.902 | 4.799     | 4.999 | 512.850            | 500.000           | 2.6   |
| Aroclor-1016-2       | 4.960 | 4.855     | 5.055 | 517.760            | 500.000           | 3.6   |
| Aroclor-1016-3       | 5.080 | 4.976     | 5.176 | 493.050            | 500.000           | -1.4  |
| Aroclor-1016-4       | 5.121 | 5.017     | 5.217 | 487.780            | 500.000           | -2.4  |
| Aroclor-1016-5       | 5.334 | 5.231     | 5.431 | 459.090            | 500.000           | -8.2  |
| Aroclor-1260-1       | 6.364 | 6.261     | 6.461 | 437.630            | 500.000           | -12.5 |
| Aroclor-1260-2       | 6.550 | 6.448     | 6.648 | 424.270            | 500.000           | -15.1 |
| Aroclor-1260-3       | 6.703 | 6.602     | 6.802 | 405.160            | 500.000           | -19.0 |
| Aroclor-1260-4       | 7.173 | 7.072     | 7.272 | 412.050            | 500.000           | -17.6 |
| Aroclor-1260-5       | 7.413 | 7.311     | 7.511 | 403.220            | 500.000           | -19.4 |
| Decachlorobiphenyl   | 8.814 | 8.715     | 8.915 | 43.980             | 50.000            | -12.0 |
| Tetrachloro-m-xylene | 3.804 | 3.699     | 3.899 | 56.760             | 50.000            | 13.5  |