

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2923

Continuing Calib Date: 08/22/2025

Initial Calibration Date(s): 08/18/2025 08/18/2025

Continuing Calib Time: 14:45

Initial Calibration Time(s): 09:53 18:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.75	4.75	4.65	4.85	0.00
Aroclor-1016-2	(2)	4.77	4.77	4.67	4.87	0.01
Aroclor-1016-3	(3)	4.82	4.82	4.72	4.92	0.00
Aroclor-1016-4	(4)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-5	(5)	5.20	5.20	5.10	5.30	0.00
Aroclor-1260-1	(1)	6.23	6.24	6.14	6.34	0.01
Aroclor-1260-2	(2)	6.42	6.43	6.33	6.53	0.01
Aroclor-1260-3	(3)	6.79	6.79	6.69	6.89	0.00
Aroclor-1260-4	(4)	7.05	7.05	6.95	7.15	0.00
Aroclor-1260-5	(5)	7.29	7.29	7.19	7.39	0.00
Tetrachloro-m-xylene		3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl		8.68	8.68	8.58	8.78	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2923

Continuing Calib Date: 08/22/2025

Initial Calibration Date(s): 08/18/2025

08/18/2025

Continuing Calib Time: 14:45

Initial Calibration Time(s): 09:53

18:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.73	4.73	4.63	4.83	0.00
Aroclor-1016-2	(2)	4.75	4.75	4.65	4.85	0.01
Aroclor-1016-3	(3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.17	5.18	5.08	5.28	0.01
Aroclor-1260-1	(1)	6.20	6.20	6.10	6.30	0.00
Aroclor-1260-2	(2)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-3	(3)	6.54	6.54	6.44	6.64	0.00
Aroclor-1260-4	(4)	7.01	7.01	6.91	7.11	0.00
Aroclor-1260-5	(5)	7.25	7.26	7.16	7.36	0.01
Tetrachloro-m-xylene		3.65	3.66	3.56	3.76	0.01
Decachlorobiphenyl		8.63	8.63	8.53	8.73	0.00