

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113148.D **Time Analyzed:** 22:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.747	4.648	4.848	543.330	500.000	8.7
Aroclor-1016-2	4.765	4.666	4.866	544.490	500.000	8.9
Aroclor-1016-3	4.823	4.724	4.924	554.330	500.000	10.9
Aroclor-1016-4	4.942	4.843	5.043	496.860	500.000	-0.6
Aroclor-1016-5	5.197	5.099	5.299	544.970	500.000	9.0
Aroclor-1260-1	6.235	6.136	6.336	516.850	500.000	3.4
Aroclor-1260-2	6.424	6.326	6.526	537.690	500.000	7.5
Aroclor-1260-3	6.790	6.691	6.891	525.940	500.000	5.2
Aroclor-1260-4	7.049	6.951	7.151	519.720	500.000	3.9
Aroclor-1260-5	7.292	7.194	7.394	513.750	500.000	2.8
Decachlorobiphenyl	8.681	8.582	8.782	50.400	50.000	0.8
Tetrachloro-m-xylene	3.663	3.563	3.763	53.800	50.000	7.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113148.D **Time Analyzed:** 22:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.728	4.629	4.829	514.570	500.000	2.9
Aroclor-1016-2	4.746	4.647	4.847	510.520	500.000	2.1
Aroclor-1016-3	4.921	4.822	5.022	500.680	500.000	0.1
Aroclor-1016-4	4.964	4.864	5.064	480.970	500.000	-3.8
Aroclor-1016-5	5.175	5.076	5.276	561.650	500.000	12.3
Aroclor-1260-1	6.203	6.104	6.304	492.350	500.000	-1.5
Aroclor-1260-2	6.391	6.292	6.492	495.230	500.000	-1.0
Aroclor-1260-3	6.543	6.443	6.643	457.930	500.000	-8.4
Aroclor-1260-4	7.011	6.912	7.112	468.030	500.000	-6.4
Aroclor-1260-5	7.254	7.155	7.355	464.900	500.000	-7.0
Decachlorobiphenyl	8.627	8.528	8.728	51.290	50.000	2.6
Tetrachloro-m-xylene	3.654	3.555	3.755	51.580	50.000	3.2