

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113193.D **Time Analyzed:** 20:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.747	4.648	4.848	518.070	500.000	3.6
Aroclor-1016-2	4.765	4.666	4.866	537.940	500.000	7.6
Aroclor-1016-3	4.823	4.724	4.924	531.100	500.000	6.2
Aroclor-1016-4	4.941	4.843	5.043	491.560	500.000	-1.7
Aroclor-1016-5	5.199	5.099	5.299	564.270	500.000	12.9
Aroclor-1260-1	6.235	6.136	6.336	480.810	500.000	-3.8
Aroclor-1260-2	6.424	6.326	6.526	473.600	500.000	-5.3
Aroclor-1260-3	6.790	6.691	6.891	462.980	500.000	-7.4
Aroclor-1260-4	7.049	6.951	7.151	482.200	500.000	-3.6
Aroclor-1260-5	7.292	7.194	7.394	465.850	500.000	-6.8
Decachlorobiphenyl	8.680	8.582	8.782	42.790	50.000	-14.4
Tetrachloro-m-xylene	3.662	3.563	3.763	52.350	50.000	4.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113193.D **Time Analyzed:** 20:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.728	4.629	4.829	494.670	500.000	-1.1
Aroclor-1016-2	4.745	4.647	4.847	501.600	500.000	0.3
Aroclor-1016-3	4.920	4.822	5.022	486.130	500.000	-2.8
Aroclor-1016-4	4.963	4.864	5.064	468.740	500.000	-6.3
Aroclor-1016-5	5.175	5.076	5.276	517.800	500.000	3.6
Aroclor-1260-1	6.202	6.104	6.304	460.660	500.000	-7.9
Aroclor-1260-2	6.390	6.292	6.492	457.400	500.000	-8.5
Aroclor-1260-3	6.542	6.443	6.643	424.220	500.000	-15.2
Aroclor-1260-4	7.011	6.912	7.112	430.870	500.000	-13.8
Aroclor-1260-5	7.253	7.155	7.355	431.340	500.000	-13.7
Decachlorobiphenyl	8.626	8.528	8.728	46.810	50.000	-6.4
Tetrachloro-m-xylene	3.655	3.555	3.755	50.400	50.000	0.8