

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2923

Continuing Calib Date: 08/23/2025

Initial Calibration Date(s): 08/18/2025 08/18/2025

Continuing Calib Time: 03:34

Initial Calibration Time(s): 09:53 18:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.75	4.75	4.65	4.85	0.00
Aroclor-1242-2 (2)	4.77	4.77	4.67	4.87	0.01
Aroclor-1242-3 (3)	4.82	4.82	4.72	4.92	0.00
Aroclor-1242-4 (4)	4.94	4.94	4.84	5.04	0.00
Aroclor-1242-5 (5)	5.59	5.59	5.49	5.69	0.00
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl	8.68	8.68	8.58	8.78	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2923

Continuing Calib Date: 08/23/2025

Initial Calibration Date(s): 08/18/2025

08/18/2025

Continuing Calib Time: 03:34

Initial Calibration Time(s): 09:53

18:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.73	4.73	4.63	4.83	0.00
Aroclor-1242-2 (2)	4.75	4.75	4.65	4.85	0.01
Aroclor-1242-3 (3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1242-4 (4)	5.00	5.01	4.91	5.11	0.01
Aroclor-1242-5 (5)	5.52	5.53	5.43	5.63	0.01
Tetrachloro-m-xylene	3.65	3.66	3.56	3.76	0.01
Decachlorobiphenyl	8.62	8.63	8.53	8.73	0.01