

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113136.D **Time Analyzed:** 17:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.748	4.648	4.848	532.940	500.000	6.6
Aroclor-1016-2	4.767	4.666	4.866	540.330	500.000	8.1
Aroclor-1016-3	4.823	4.724	4.924	537.490	500.000	7.5
Aroclor-1016-4	4.943	4.843	5.043	497.460	500.000	-0.5
Aroclor-1016-5	5.199	5.099	5.299	538.190	500.000	7.6
Aroclor-1260-1	6.236	6.136	6.336	501.250	500.000	0.3
Aroclor-1260-2	6.426	6.326	6.526	502.220	500.000	0.4
Aroclor-1260-3	6.791	6.691	6.891	492.210	500.000	-1.6
Aroclor-1260-4	7.051	6.951	7.151	485.430	500.000	-2.9
Aroclor-1260-5	7.294	7.194	7.394	489.850	500.000	-2.0
Decachlorobiphenyl	8.682	8.582	8.782	47.850	50.000	-4.3
Tetrachloro-m-xylene	3.664	3.563	3.763	53.240	50.000	6.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113136.D **Time Analyzed:** 17:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.729	4.629	4.829	506.020	500.000	1.2
Aroclor-1016-2	4.747	4.647	4.847	506.130	500.000	1.2
Aroclor-1016-3	4.922	4.822	5.022	491.760	500.000	-1.6
Aroclor-1016-4	4.964	4.864	5.064	476.220	500.000	-4.8
Aroclor-1016-5	5.176	5.076	5.276	513.300	500.000	2.7
Aroclor-1260-1	6.204	6.104	6.304	468.970	500.000	-6.2
Aroclor-1260-2	6.392	6.292	6.492	466.240	500.000	-6.8
Aroclor-1260-3	6.543	6.443	6.643	439.000	500.000	-12.2
Aroclor-1260-4	7.012	6.912	7.112	446.520	500.000	-10.7
Aroclor-1260-5	7.255	7.155	7.355	444.400	500.000	-11.1
Decachlorobiphenyl	8.628	8.528	8.728	49.290	50.000	-1.4
Tetrachloro-m-xylene	3.655	3.555	3.755	51.320	50.000	2.6