

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL07 **Date Analyzed:** 09/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113443.D **Time Analyzed:** 19:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.744	4.648	4.848	549.920	500.000	10.0
Aroclor-1016-2	4.763	4.666	4.866	553.020	500.000	10.6
Aroclor-1016-3	4.820	4.724	4.924	550.340	500.000	10.1
Aroclor-1016-4	4.938	4.843	5.043	498.420	500.000	-0.3
Aroclor-1016-5	5.196	5.099	5.299	589.060	500.000	17.8
Aroclor-1260-1	6.232	6.136	6.336	512.540	500.000	2.5
Aroclor-1260-2	6.422	6.326	6.526	507.110	500.000	1.4
Aroclor-1260-3	6.787	6.691	6.891	487.580	500.000	-2.5
Aroclor-1260-4	7.046	6.951	7.151	490.580	500.000	-1.9
Aroclor-1260-5	7.289	7.194	7.394	468.670	500.000	-6.3
Decachlorobiphenyl	8.676	8.582	8.782	42.670	50.000	-14.7
Tetrachloro-m-xylene	3.661	3.563	3.763	54.990	50.000	10.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL07 **Date Analyzed:** 09/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113443.D **Time Analyzed:** 19:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.725	4.629	4.829	527.550	500.000	5.5
Aroclor-1016-2	4.744	4.647	4.847	519.260	500.000	3.9
Aroclor-1016-3	4.918	4.822	5.022	501.870	500.000	0.4
Aroclor-1016-4	4.961	4.864	5.064	483.460	500.000	-3.3
Aroclor-1016-5	5.173	5.076	5.276	550.660	500.000	10.1
Aroclor-1260-1	6.200	6.104	6.304	472.290	500.000	-5.5
Aroclor-1260-2	6.388	6.292	6.492	468.570	500.000	-6.3
Aroclor-1260-3	6.539	6.443	6.643	427.700	500.000	-14.5
Aroclor-1260-4	7.008	6.912	7.112	433.710	500.000	-13.3
Aroclor-1260-5	7.251	7.155	7.355	424.440	500.000	-15.1
Decachlorobiphenyl	8.622	8.528	8.728	49.320	50.000	-1.4
Tetrachloro-m-xylene	3.653	3.555	3.755	53.430	50.000	6.9