

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL09 **Date Analyzed:** 08/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074556.D **Time Analyzed:** 19:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.811	5.706	5.906	476.060	500.000	-4.8
Aroclor-1016-2	5.833	5.727	5.927	493.460	500.000	-1.3
Aroclor-1016-3	5.896	5.790	5.990	453.750	500.000	-9.3
Aroclor-1016-4	5.993	5.888	6.088	483.490	500.000	-3.3
Aroclor-1016-5	6.285	6.180	6.380	501.400	500.000	0.3
Aroclor-1260-1	7.403	7.297	7.497	477.480	500.000	-4.5
Aroclor-1260-2	7.656	7.550	7.750	480.720	500.000	-3.9
Aroclor-1260-3	8.013	7.908	8.108	514.970	500.000	3.0
Aroclor-1260-4	8.242	8.136	8.336	497.970	500.000	-0.4
Aroclor-1260-5	8.568	8.462	8.662	481.020	500.000	-3.8
Decachlorobiphenyl	10.439	10.333	10.533	48.860	50.000	-2.3
Tetrachloro-m-xylene	4.660	4.554	4.754	48.040	50.000	-3.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL09 **Date Analyzed:** 08/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074556.D **Time Analyzed:** 19:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.902	4.799	4.999	497.550	500.000	-0.5
Aroclor-1016-2	4.959	4.855	5.055	505.540	500.000	1.1
Aroclor-1016-3	5.079	4.976	5.176	502.390	500.000	0.5
Aroclor-1016-4	5.120	5.017	5.217	478.010	500.000	-4.4
Aroclor-1016-5	5.335	5.231	5.431	449.060	500.000	-10.2
Aroclor-1260-1	6.364	6.261	6.461	466.850	500.000	-6.6
Aroclor-1260-2	6.550	6.448	6.648	451.180	500.000	-9.8
Aroclor-1260-3	6.704	6.602	6.802	418.490	500.000	-16.3
Aroclor-1260-4	7.174	7.072	7.272	452.890	500.000	-9.4
Aroclor-1260-5	7.413	7.311	7.511	441.700	500.000	-11.7
Decachlorobiphenyl	8.816	8.715	8.915	44.520	50.000	-11.0
Tetrachloro-m-xylene	3.802	3.699	3.899	50.270	50.000	0.5