

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL11 **Date Analyzed:** 08/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074577.D **Time Analyzed:** 14:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.811	5.706	5.906	476.170	500.000	-4.8
Aroclor-1016-2	5.832	5.727	5.927	501.970	500.000	0.4
Aroclor-1016-3	5.896	5.790	5.990	461.020	500.000	-7.8
Aroclor-1016-4	5.993	5.888	6.088	487.570	500.000	-2.5
Aroclor-1016-5	6.285	6.180	6.380	507.720	500.000	1.5
Aroclor-1260-1	7.403	7.297	7.497	487.310	500.000	-2.5
Aroclor-1260-2	7.655	7.550	7.750	478.880	500.000	-4.2
Aroclor-1260-3	8.013	7.908	8.108	507.660	500.000	1.5
Aroclor-1260-4	8.242	8.136	8.336	489.410	500.000	-2.1
Aroclor-1260-5	8.568	8.462	8.662	483.040	500.000	-3.4
Decachlorobiphenyl	10.440	10.333	10.533	50.020	50.000	0.0
Tetrachloro-m-xylene	4.659	4.554	4.754	48.500	50.000	-3.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL11 **Date Analyzed:** 08/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074577.D **Time Analyzed:** 14:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.901	4.799	4.999	494.070	500.000	-1.2
Aroclor-1016-2	4.959	4.855	5.055	509.360	500.000	1.9
Aroclor-1016-3	5.078	4.976	5.176	495.350	500.000	-0.9
Aroclor-1016-4	5.119	5.017	5.217	495.570	500.000	-0.9
Aroclor-1016-5	5.334	5.231	5.431	462.950	500.000	-7.4
Aroclor-1260-1	6.364	6.261	6.461	467.730	500.000	-6.5
Aroclor-1260-2	6.551	6.448	6.648	460.480	500.000	-7.9
Aroclor-1260-3	6.705	6.602	6.802	428.120	500.000	-14.4
Aroclor-1260-4	7.174	7.072	7.272	461.240	500.000	-7.8
Aroclor-1260-5	7.413	7.311	7.511	463.780	500.000	-7.2
Decachlorobiphenyl	8.817	8.715	8.915	45.240	50.000	-9.5
Tetrachloro-m-xylene	3.802	3.699	3.899	49.780	50.000	-0.4