

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL14 **Date Analyzed:** 09/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074819.D **Time Analyzed:** 16:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.813	5.706	5.906	503.390	500.000	0.7
Aroclor-1016-2	5.834	5.727	5.927	519.210	500.000	3.8
Aroclor-1016-3	5.897	5.790	5.990	470.950	500.000	-5.8
Aroclor-1016-4	5.994	5.888	6.088	497.160	500.000	-0.6
Aroclor-1016-5	6.287	6.180	6.380	515.490	500.000	3.1
Aroclor-1260-1	7.404	7.297	7.497	474.610	500.000	-5.1
Aroclor-1260-2	7.657	7.550	7.750	462.930	500.000	-7.4
Aroclor-1260-3	8.015	7.908	8.108	487.570	500.000	-2.5
Aroclor-1260-4	8.243	8.136	8.336	460.910	500.000	-7.8
Aroclor-1260-5	8.569	8.462	8.662	446.490	500.000	-10.7
Decachlorobiphenyl	10.440	10.333	10.533	54.250	50.000	8.5
Tetrachloro-m-xylene	4.661	4.554	4.754	52.870	50.000	5.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL14 **Date Analyzed:** 09/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074819.D **Time Analyzed:** 16:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.902	4.799	4.999	512.850	500.000	2.6
Aroclor-1016-2	4.960	4.855	5.055	517.760	500.000	3.6
Aroclor-1016-3	5.080	4.976	5.176	493.050	500.000	-1.4
Aroclor-1016-4	5.121	5.017	5.217	487.780	500.000	-2.4
Aroclor-1016-5	5.334	5.231	5.431	459.090	500.000	-8.2
Aroclor-1260-1	6.364	6.261	6.461	437.630	500.000	-12.5
Aroclor-1260-2	6.550	6.448	6.648	424.270	500.000	-15.1
Aroclor-1260-3	6.703	6.602	6.802	405.160	500.000	-19.0
Aroclor-1260-4	7.173	7.072	7.272	412.050	500.000	-17.6
Aroclor-1260-5	7.413	7.311	7.511	403.220	500.000	-19.4
Decachlorobiphenyl	8.814	8.715	8.915	43.980	50.000	-12.0
Tetrachloro-m-xylene	3.804	3.699	3.899	56.760	50.000	13.5