

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL15 **Date Analyzed:** 09/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074833.D **Time Analyzed:** 21:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.811	5.706	5.906	511.910	500.000	2.4
Aroclor-1016-2	5.833	5.727	5.927	530.710	500.000	6.1
Aroclor-1016-3	5.896	5.790	5.990	481.440	500.000	-3.7
Aroclor-1016-4	5.993	5.888	6.088	508.490	500.000	1.7
Aroclor-1016-5	6.285	6.180	6.380	532.310	500.000	6.5
Aroclor-1260-1	7.403	7.297	7.497	484.040	500.000	-3.2
Aroclor-1260-2	7.655	7.550	7.750	479.170	500.000	-4.2
Aroclor-1260-3	8.013	7.908	8.108	498.430	500.000	-0.3
Aroclor-1260-4	8.241	8.136	8.336	478.630	500.000	-4.3
Aroclor-1260-5	8.567	8.462	8.662	460.000	500.000	-8.0
Decachlorobiphenyl	10.438	10.333	10.533	55.560	50.000	11.1
Tetrachloro-m-xylene	4.660	4.554	4.754	53.530	50.000	7.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2923
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL15 **Date Analyzed:** 09/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074833.D **Time Analyzed:** 21:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.900	4.799	4.999	525.140	500.000	5.0
Aroclor-1016-2	4.958	4.855	5.055	541.260	500.000	8.3
Aroclor-1016-3	5.078	4.976	5.176	511.240	500.000	2.2
Aroclor-1016-4	5.119	5.017	5.217	497.660	500.000	-0.5
Aroclor-1016-5	5.333	5.231	5.431	472.230	500.000	-5.6
Aroclor-1260-1	6.362	6.261	6.461	466.130	500.000	-6.8
Aroclor-1260-2	6.549	6.448	6.648	457.120	500.000	-8.6
Aroclor-1260-3	6.703	6.602	6.802	423.090	500.000	-15.4
Aroclor-1260-4	7.172	7.072	7.272	448.410	500.000	-10.3
Aroclor-1260-5	7.412	7.311	7.511	437.440	500.000	-12.5
Decachlorobiphenyl	8.813	8.715	8.915	47.780	50.000	-4.4
Tetrachloro-m-xylene	3.802	3.699	3.899	58.730	50.000	17.5