



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q2923	Analytical Method:	8082A
Client:	ATC Group Services LLC	Datafile :	PP074574.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB169353BS (Column 1)	AR1016	166.5	142	ug/kg	85				71	120	
	AR1260	166.5	137	ug/kg	82				65	130	
PB169353BS (Column 2)	AR1016	166.5	124	ug/kg	74				71	120	
	AR1260	166.5	125	ug/kg	75				65	130	