

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2924

Analytical Method: 8270E

Client: AECOM

DataFile: BP025581.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169362BS	Benzaldehyde	50	22.9	ug/L	46				10	162	
	Phenol	50	41.7	ug/L	83				66	118	
	bis(2-Chloroethyl)ether	50	40.1	ug/L	80				62	103	
	2-Chlorophenol	50	41.6	ug/L	83				70	117	
	2-Methylphenol	50	41.0	ug/L	82				69	109	
	2,2-oxybis(1-Chloropropane)	50	39.7	ug/L	79				65	100	
	Acetophenone	50	39.8	ug/L	80				60	104	
	3+4-Methylphenols	50	39.8	ug/L	80				67	106	
	N-Nitroso-di-n-propylamine	50	36.5	ug/L	73				57	107	
	Hexachloroethane	50	40.3	ug/L	81				76	118	
	Nitrobenzene	50	41.1	ug/L	82				58	106	
	Isophorone	50	39.2	ug/L	78				61	102	
	2-Nitrophenol	50	42.7	ug/L	85				70	115	
	2,4-Dimethylphenol	50	42.3	ug/L	85				42	142	
	bis(2-Chloroethoxy)methane	50	40.1	ug/L	80				58	109	
	2,4-Dichlorophenol	50	43.3	ug/L	87				66	115	
	Naphthalene	50	39.7	ug/L	79				64	107	
	4-Chloroaniline	50	23.9	ug/L	48				10	85	
	Hexachlorobutadiene	50	40.8	ug/L	82				69	101	
	Caprolactam	50	39.6	ug/L	79				58	128	
	4-Chloro-3-methylphenol	50	42.3	ug/L	85				65	114	
	2-Methylnaphthalene	50	39.4	ug/L	79				64	107	
	Hexachlorocyclopentadiene	100	81.0	ug/L	81				36	160	
	2,4,6-Trichlorophenol	50	44.1	ug/L	88				61	110	
	2,4,5-Trichlorophenol	50	44.6	ug/L	89				70	106	
	1,1-Biphenyl	50	41.8	ug/L	84				72	98	
	2-Chloronaphthalene	50	40.9	ug/L	82				59	106	
	2-Nitroaniline	50	44.0	ug/L	88				73	114	
	Dimethylphthalate	50	41.0	ug/L	82				64	103	
	Acenaphthylene	50	40.9	ug/L	82				79	103	
	2,6-Dinitrotoluene	50	43.4	ug/L	87				64	110	
	3-Nitroaniline	50	31.2	ug/L	62				28	100	
	Acenaphthene	50	39.5	ug/L	79				59	113	
	2,4-Dinitrophenol	100	100	ug/L	100				36	166	
	4-Nitrophenol	100	82.1	ug/L	82				45	147	
	Dibenzofuran	50	40.2	ug/L	80				65	106	
	2,4-Dinitrotoluene	50	43.7	ug/L	87				60	115	
	Diethylphthalate	50	41.5	ug/L	83				63	105	
	4-Chlorophenyl-phenylether	50	40.1	ug/L	80				61	104	
	Fluorene	50	40.2	ug/L	80				64	107	
	4-Nitroaniline	50	41.1	ug/L	82				55	125	
	4,6-Dinitro-2-methylphenol	50	46.7	ug/L	93				62	132	
	N-Nitrosodiphenylamine	50	42.8	ug/L	86				61	109	
	4-Bromophenyl-phenylether	50	42.2	ug/L	84				73	103	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2924

Analytical Method: 8270E

Client: AECOM

DataFile: BP025581.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	RPD		Limits		RPD
							Qual	Qual	Low	High	
PB169362BS	Hexachlorobenzene	50	42.0	ug/L	84				73	106	
	Atrazine	50	41.7	ug/L	83				76	120	
	Pentachlorophenol	100	86.9	ug/L	87				47	114	
	Phenanthrene	50	40.9	ug/L	82				62	109	
	Anthracene	50	41.4	ug/L	83				65	110	
	Carbazole	50	40.8	ug/L	82				62	106	
	Di-n-butylphthalate	50	41.6	ug/L	83				64	106	
	Fluoranthene	50	40.0	ug/L	80				64	110	
	Pyrene	50	43.4	ug/L	87				71	103	
	Butylbenzylphthalate	50	44.3	ug/L	89				61	105	
	3,3-Dichlorobenzidine	50	31.9	ug/L	64				43	108	
	Benzo(a)anthracene	50	42.2	ug/L	84				62	107	
	Chrysene	50	41.7	ug/L	83				61	108	
	bis(2-Ethylhexyl)phthalate	50	42.7	ug/L	85				59	110	
	Di-n-octyl phthalate	50	43.7	ug/L	87				52	139	
	Benzo(b)fluoranthene	50	41.7	ug/L	83				77	113	
	Benzo(k)fluoranthene	50	41.8	ug/L	84				77	105	
	Benzo(a)pyrene	50	41.6	ug/L	83				72	131	
	Indeno(1,2,3-cd)pyrene	50	41.4	ug/L	83				72	105	
	Dibenz(a,h)anthracene	50	41.8	ug/L	84				78	115	
	Benzo(g,h,i)perylene	50	41.4	ug/L	83				75	118	
	1,2,4,5-Tetrachlorobenzene	50	42.0	ug/L	84				72	101	
	1,4-Dioxane	50	34.9	ug/L	70				38	125	
	2,3,4,6-Tetrachlorophenol	50	43.0	ug/L	86				63	116	