

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2924

Client ID : SSTDCCC040

Date Analyzed: 08/22/2025

Lab File ID: BF143563.D

Time Analyzed: 10:29

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	94271	6.928	366222	8.21	205977	9.96
UPPER LIMIT	188542	7.428	732444	8.71	411954	10.463
LOWER LIMIT	47135.5	6.428	183111	7.71	102989	9.463
EPA SAMPLE NO.						
01 MW-10C-082025	104622	6.92	416287	8.20	252150	9.96
02 MW-2B-082025	97960	6.93	310044	8.23	220938	9.96

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.