

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2924

Client ID : SSTDCCC040

Date Analyzed: 08/27/2025

Lab File ID: BF143596.D

Time Analyzed: 09:16

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	115923	6.893	446099	8.18	249607	9.93
UPPER LIMIT	231846	7.393	892198	8.675	499214	10.428
LOWER LIMIT	57961.5	6.393	223050	7.675	124804	9.428
EPA SAMPLE NO.						
01 MW-18C-082025	107445	6.89	403245	8.17	213409	9.92
02 FB-01-082025	111189	6.89	428717	8.17	235375	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.