

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2924

Client ID : SSTDCCC040

Date Analyzed: 08/26/2025

Lab File ID: BP025579.D

Time Analyzed: 10:21

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	235051	7.778	932609	10.55	572065	14.39
UPPER LIMIT	470102	8.278	1865220	11.049	1144130	14.89
LOWER LIMIT	117526	7.278	466305	10.049	286033	13.89
EPA SAMPLE NO.						
01 PB169362BS	256194	7.78	1004580	10.55	636349	14.41
02 MW-2B-082025DL	167152	7.77	645100	10.55	403608	14.39
03 MW-2B-082025DL2	173400	7.78	700981	10.54	444336	14.39
04 MW-2A-082025	209304	7.78	834432	10.54	540709	14.38
05 MW-16A-082025	194953	7.78	750992	10.55	492017	14.39
06 MW-16C-082025	200652	7.77	790738	10.55	502369	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.