

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance  
Lab Code: ACE  
Lab File ID: BF143595.D  
Instrument ID: BNA\_F

Contract: AECO02  
SDG NO.: Q2924  
DFTPP Injection Date: 08/27/2025  
DFTPP Injection Time: 08:48

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 68  | Less than 2.0% of mass 69          | 0.6 ( 1.6 ) 1        |
| 69  | Mass 69 relative abundance         | 33.7                 |
| 70  | Less than 2.0% of mass 69          | 0.2 ( 0.6 ) 1        |
| 197 | Less than 2.0% of mass 198         | 0.0                  |
| 198 | Base Peak, 100% relative abundance | 100                  |
| 199 | 5.0 to 9.0% of mass 198            | 6.6                  |
| 365 | Greater than 1% of mass 198        | 3.2                  |
| 441 | Present, but less than mass 443    | 16                   |
| 442 | Greater than 50% of mass 198       | 100                  |
| 443 | 15.0 - 24.0% of mass 442           | 19.2 ( 19.2 ) 2      |

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED | TIME<br>ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| SSTDCCC040        | SSTDCCC040       | BF143596.D     | 08/27/2025       | 09:16            |
| FB-01-082025      | Q2924-11         | BF143601.D     | 08/27/2025       | 11:53            |
| MW-18C-082025     | Q2924-07         | BF143603.D     | 08/27/2025       | 12:52            |