

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP025595.D
Instrument ID: BNA_P

Contract: AEC002
SDG NO.: Q2924
DFTPP Injection Date: 08/27/2025
DFTPP Injection Time: 08:58

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.7) 1
69	Mass 69 relative abundance	31.5
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	15.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.6 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP025596.D	08/27/2025	09:39
MW-201-082025	Q2924-02	BP025601.D	08/27/2025	13:37
MW-201-082025MS	Q2924-03MS	BP025602.D	08/27/2025	14:18
MW-201-082025MSD	Q2924-04MSD	BP025603.D	08/27/2025	15:00
MW-201-082025DL	Q2924-02DL	BP025607.D	08/27/2025	17:44
MW-201-082025DL2	Q2924-02DL2	BP025608.D	08/27/2025	18:26