



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: AECO02
Lab Code: ACE SDG No.: Q2924
Instrument ID: MSVOA_V Calibration Date/Time: 08/21/2025 15:42
Lab File ID: VV038982.D Init. Calib. Date(s): 08/21/2025 08/21/2025
Heated Purge: (Y/N) N Init. Calib. Time(s): 09:05 11:45
GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Bromochloromethane	50.000	56.97	-13.94	20
2-Hexanone	250.000	224.81	10.078	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.