

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG No.: Q2924
 Instrument ID: MSVOA_V Calibration Date/Time: 08/22/2025 09:03
 Lab File ID: VV039009.D Init. Calib. Date(s): 08/21/2025 08/21/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 09:05 11:45
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Bromochloromethane	50.000	58.41	-16.818	20
2-Hexanone	250.000	218.89	12.4448	20