

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2924</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>AECOM</u>	Datafile :	<u>VV038985.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VV0821WBS01	Dichlorodifluoromethane	20	17.3	ug/L	86			69	116	
	Chloromethane	20	15.8	ug/L	79			65	116	
	Vinyl chloride	20	16.0	ug/L	80			65	117	
	Bromomethane	20	19.0	ug/L	95			58	125	
	Chloroethane	20	19.4	ug/L	97			56	128	
	Trichlorofluoromethane	20	19.0	ug/L	95			73	115	
	1,1,2-Trichlorotrifluoroethane	20	18.6	ug/L	93			80	112	
	1,1-Dichloroethene	20	19.2	ug/L	96			74	110	
	Acetone	100	94.1	ug/L	94			60	125	
	Carbon disulfide	20	19.2	ug/L	96			64	112	
	Methyl tert-butyl Ether	20	20.3	ug/L	102			78	114	
	Methyl Acetate	20	20.6	ug/L	103			67	125	
	Methylene Chloride	20	18.7	ug/L	94			72	114	
	trans-1,2-Dichloroethene	20	18.9	ug/L	95			75	108	
	1,1-Dichloroethane	20	17.8	ug/L	89			78	112	
	Cyclohexane	20	18.2	ug/L	91			75	110	
	2-Butanone	100	94.3	ug/L	94			65	122	
	Carbon Tetrachloride	20	16.5	ug/L	83			77	113	
	cis-1,2-Dichloroethene	20	17.9	ug/L	90			77	110	
	Bromochloromethane	20	16.0	ug/L	80			70	124	
	Chloroform	20	18.1	ug/L	91			79	113	
	1,1,1-Trichloroethane	20	17.6	ug/L	88			80	108	
	Methylcyclohexane	20	18.7	ug/L	94			72	115	
	Benzene	20	19.2	ug/L	96			82	109	
	1,2-Dichloroethane	20	18.4	ug/L	92			80	115	
	Trichloroethene	20	18.0	ug/L	90			77	113	
	1,2-Dichloropropane	20	19.4	ug/L	97			83	111	
	Bromodichloromethane	20	18.1	ug/L	91			83	110	
	4-Methyl-2-Pentanone	100	110	ug/L	110			74	118	
	Toluene	20	19.4	ug/L	97			82	110	
	t-1,3-Dichloropropene	20	17.9	ug/L	90			79	110	
	cis-1,3-Dichloropropene	20	19.5	ug/L	98			82	110	
	1,1,2-Trichloroethane	20	17.6	ug/L	88			83	112	
	2-Hexanone	100	82.7	ug/L	83			73	117	
	Dibromochloromethane	20	17.2	ug/L	86			82	110	
	1,2-Dibromoethane	20	18.3	ug/L	92			81	110	
	Tetrachloroethene	20	18.6	ug/L	93			67	123	
	Chlorobenzene	20	18.0	ug/L	90			82	109	
	Ethyl Benzene	20	19.1	ug/L	96			83	109	
	m/p-Xylenes	40	36.4	ug/L	91			82	110	
	o-Xylene	20	19.6	ug/L	98			83	109	
	Styrene	20	20.4	ug/L	102			80	111	
	Bromoform	20	18.6	ug/L	93			79	109	
	Isopropylbenzene	20	19.7	ug/L	99			83	112	
	1,1,2,2-Tetrachloroethane	20	17.8	ug/L	89			76	118	
	1,3-Dichlorobenzene	20	18.6	ug/L	93			82	108	
	1,4-Dichlorobenzene	20	18.8	ug/L	94			82	107	
	1,2-Dichlorobenzene	20	18.6	ug/L	93			82	109	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2924 Analytical Method: SW8260-Low
Client: AECOM Datafile : VV038985.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VV0821WBS01	1,2-Dibromo-3-Chloropropane	20	19.1	ug/L	96			68	112	
	1,2,4-Trichlorobenzene	20	17.0	ug/L	85			75	113	
	1,2,3-Trichlorobenzene	20	18.4	ug/L	92			76	114	