

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q2924
 Lab File ID: VV039009.D Date Analyzed: 08/22/2025
 Instrument ID: MSVOA_V Time Analyzed: 09:03
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	735484	4.60	1263800	5.53	1133330	8.77
UPPER LIMIT	1470970	5.097	2527600	6.029	2266660	9.273
LOWER LIMIT	367742	4.097	631899	5.029	566665	8.273
EPA SAMPLE NO.						
MW-201-082025DL	581936	4.60	1169048	5.53	1041283	8.78
MW-201-082025DL2	766342	4.60	1437458	5.53	1305700	8.77
MW-2B-082025DL	617177	4.60	1163830	5.53	1033790	8.78
MW-2B-082025DL2	582933	4.60	1168109	5.53	1042248	8.77
VV0822WBL01	530331	4.60	1082642	5.53	991839	8.78
VV0822WBS01	711068	4.60	1223613	5.53	1111109	8.77

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.