

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2924</u>
Lab File ID:	<u>VV038982.D</u>	Date Analyzed:	<u>08/21/2025</u>
Instrument ID:	<u>MSVOA_V</u>	Time Analyzed:	<u>15:42</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	452177	11.172				
UPPER LIMIT	904354	11.672				
LOWER LIMIT	226089	10.672				
EPA SAMPLE NO.						
MW-10C-082025	354625	11.18				
MW-201-082025	442841	11.18				
MW-201-082025MS	531973	11.18				
MW-201-082025MSD	478184	11.18				
MW-2B-082025	411561	11.18				
MW-2A-082025	448609	11.18				
MW-18C-082025	449215	11.18				
MW-16A-082025	493806	11.18				
MW-16C-082025	439427	11.17				
MW-8A-082025	375249	11.18				
FB-01-082025	424569	11.18				
TB	415869	11.18				
VV0821WBL01	318475	11.18				
VV0821WBS01	439751	11.17				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.