

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2924</u>
Lab File ID:	<u>VV039009.D</u>	Date Analyzed:	<u>08/22/2025</u>
Instrument ID:	<u>MSVOA_V</u>	Time Analyzed:	<u>09:03</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	587247	11.172				
UPPER LIMIT	1174490	11.672				
LOWER LIMIT	293624	10.672				
EPA SAMPLE NO.						
MW-201-082025DL	510275	11.17				
MW-201-082025DL2	617094	11.17				
MW-2B-082025DL	477379	11.17				
MW-2B-082025DL2	457197	11.17				
VV0822WBL01	416909	11.18				
VV0822WBS01	587275	11.17				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.