

Initial and Continuing Calibration Verification

Client: ATG-GREENVILLE AEC

SDG No.: Q2925

Project: AEC-2025-0013- Durham School Services - Minetto NY

RunNo.: LB136918

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	50.962	50	102	95-105	05/28/2025
Sample ID: CCV1 COD	mg/L	49.946	50	100	95-105	08/22/2025
Sample ID: CCV2 COD	mg/L	48.931	50	98	95-105	08/22/2025
Sample ID: CCV3 COD	mg/L	49.946	50	100	95-105	08/22/2025