

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2925
Project:	AEC-2025-0013- Durham School Services - Minetto N	Sample ID:	Q2925-01
Client ID:	OUTFALL-2MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	21.4		1.30	J	20.0	1	101		08/21/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2925
Project:	AEC-2025-0013- Durham School Services - Minetto N	Sample ID:	Q2925-01
Client ID:	OUTFALL-2MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	21.1		1.30	J	20.0	1	99		08/21/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2925
Project:	AEC-2025-0013- Durham School Services - Minetto N	Sample ID:	Q2931-01
Client ID:	DA1-1MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	55.0		13.4		50.0	1	83		08/22/2025

Matrix Spike Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2925
Project:	AEC-2025-0013- Durham School Services - Minetto N	Sample ID:	Q2931-01
Client ID:	DA1-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
COD	mg/L	75-125	52.0		13.4		50.0	1	77		08/22/2025