

Duplicate Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2925
Project:	AEC-2025-0013- Durham School Services - Minetto NY	Sample ID:	Q2925-01
Client ID:	OUTFALL-2MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	21.4		21.1		1	1.41		08/21/2025

Duplicate Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2925
Project:	AEC-2025-0013- Durham School Services - Minetto NY	Sample ID:	Q2931-01
Client ID:	DA1-1DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	13.4		13.4		1	0		08/22/2025

Duplicate Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2925
Project:	AEC-2025-0013- Durham School Services - Minetto NY	Sample ID:	Q2931-01
Client ID:	DA1-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
COD	mg/L	+/-20	55.0		52.0		1	5.61		08/22/2025