

Laboratory Control Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2925
Project:	AEC-2025-0013- Durham School Services - Minetto NY	Run No.:	LB136909

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136909BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	08/21/2025

Laboratory Control Sample Summary

Client:	ATG-GREENVILLE AEC	SDG No.:	Q2925
Project:	AEC-2025-0013- Durham School Services - Minetto NY	Run No.:	LB136918

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136918BS							
COD	mg/L	50	49.9		100	1	90-110	08/22/2025