

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2925</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ATG-GREENVILLE AEC</u>	Datafile :	<u>VV038985.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VV0821WBS01	Benzene	20	19.2	ug/L	96			82	109	
	Toluene	20	19.4	ug/L	97			82	110	
	Ethyl Benzene	20	19.1	ug/L	96			83	109	
	m/p-Xylenes	40	36.4	ug/L	91			82	110	
	o-Xylene	20	19.6	ug/L	98			83	109	