

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ATGG01
 Lab Code: ACE SDG NO.: Q2925
 Lab File ID: VV038982.D Date Analyzed: 08/21/2025
 Instrument ID: MSVOA_V Time Analyzed: 15:42
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	452177	11.172				
UPPER LIMIT	904354	11.672				
LOWER LIMIT	226089	10.672				
EPA SAMPLE NO.						
OUTFALL-2	375523	11.18				
VV0821WBL01	318475	11.18				
VV0821WBS01	439751	11.17				
VV0821WBSD01	414371	11.17				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.