



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Hero Construction

Project Name: Residential Soil Testing

Project # N/A

Order ID # Q2928

Test Name: Cyanide

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 08/21/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Cyanide. This data package contains results for Cyanide.

C. Analytical Techniques:

The analysis of Cyanide was based on method 9012B.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike analysis met criteria for all compounds.

The Matrix Spike Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____