

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Hero Construction

Project Name: Residential Soil Testing

Project # N/A Order ID # O2928

Test Name: Mercury, Metals ICP-TAL

A. Number of Samples and Date of Receipt:

2 Solid samples were received on 08/21/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Cyanide, EPH_F2, Mercury, Metals ICP-TAL, METALS-TAL, PCB, Pesticide-TCL, SVOC-TCL BNA -20 and VOC-TCLVOA-10. This data package contains results for Mercury, Metals ICP-TAL.

C. Analytical Techniques:

The analysis of Metals ICP-TAL was based on method 6010D, digestion based on method 3050 (soils). The analysis and digestion of Mercury was based on method 7471B.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

Sample 27-BELL was diluted due to high concentrations for Mercury.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike (29-SEPTICMS) analysis met criteria for all compounds except for Antimony and Chromium due to Chemical Interference during Digestion process.

The Matrix Spike Duplicate (29-SEPTICMSD) analysis met criteria for all compounds except for Antimony and Thallium due to Chemical Interference during Digestion process.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature			
Signature_			