



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Hero Construction

SDG No.: Q2928

Project: Residential Soil Testing

RunNo.: LB136947

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.1	0.099	101	90-110	08/25/2025
Sample ID: Cyanide	CCV1	mg/L	0.25	0.25	100	90-110	08/25/2025
Sample ID: Cyanide	CCV2	mg/L	0.25	0.25	100	90-110	08/25/2025