

Duplicate Sample Summary

Client:	Hero Construction	SDG No.:	Q2928
Project:	Residential Soil Testing	Sample ID:	Q2932-03
Client ID:	ARS20-0002-SOIDUP	Percent Solids for Spike Sample:	95.5

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	0.044	U	0.043	U	1	0		08/25/2025

Duplicate Sample Summary

Client:	Hero Construction	SDG No.:	Q2928
Project:	Residential Soil Testing	Sample ID:	Q2932-03
Client ID:	ARS20-0002-SOILMSD	Percent Solids for Spike Sample:	95.5

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	1.90		2.00		1	5		08/25/2025