

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HERO01
Lab Code: ACE **SDG NO.:** Q2928
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074592.D **Time Analyzed:** 19:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.809	5.706	5.906	478.870	500.000	-4.2
Aroclor-1016-2	5.831	5.727	5.927	494.830	500.000	-1.0
Aroclor-1016-3	5.895	5.790	5.990	462.970	500.000	-7.4
Aroclor-1016-4	5.991	5.888	6.088	482.320	500.000	-3.5
Aroclor-1016-5	6.283	6.180	6.380	498.280	500.000	-0.3
Aroclor-1260-1	7.401	7.297	7.497	479.010	500.000	-4.2
Aroclor-1260-2	7.653	7.550	7.750	469.790	500.000	-6.0
Aroclor-1260-3	8.011	7.908	8.108	497.950	500.000	-0.4
Aroclor-1260-4	8.240	8.136	8.336	480.450	500.000	-3.9
Aroclor-1260-5	8.566	8.462	8.662	474.990	500.000	-5.0
Decachlorobiphenyl	10.435	10.333	10.533	48.800	50.000	-2.4
Tetrachloro-m-xylene	4.659	4.554	4.754	48.220	50.000	-3.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** HERO01
Lab Code: ACE **SDG NO.:** Q2928
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074592.D **Time Analyzed:** 19:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.901	4.799	4.999	488.860	500.000	-2.2
Aroclor-1016-2	4.958	4.855	5.055	504.890	500.000	1.0
Aroclor-1016-3	5.079	4.976	5.176	481.500	500.000	-3.7
Aroclor-1016-4	5.119	5.017	5.217	480.570	500.000	-3.9
Aroclor-1016-5	5.333	5.231	5.431	468.870	500.000	-6.2
Aroclor-1260-1	6.362	6.261	6.461	468.040	500.000	-6.4
Aroclor-1260-2	6.549	6.448	6.648	450.670	500.000	-9.9
Aroclor-1260-3	6.703	6.602	6.802	414.420	500.000	-17.1
Aroclor-1260-4	7.172	7.072	7.272	450.160	500.000	-10.0
Aroclor-1260-5	7.412	7.311	7.511	441.440	500.000	-11.7
Decachlorobiphenyl	8.814	8.715	8.915	44.310	50.000	-11.4
Tetrachloro-m-xylene	3.802	3.699	3.899	50.520	50.000	1.0